



UNIVERSITI PUTRA MALAYSIA

**MINIMIZATION OF ENTRANCE SURFACE DOSE AND CRITICAL
ORGAN DOSE FOR MEDICAL RADIOGRAPHY USING
OPTIMIZATION PROCEDURES**

**LOTF ALI MEHDIPOUR NASSAB RABOR
FS 2010 2**





**MINIMIZATION OF ENTRANCE SURFACE
DOSE AND CRITICAL ORGAN DOSE FOR
MEDICAL RADIOGRAPHY USING
OPTIMIZATION PROCEDURES**

LOTF ALI MEHDIPOUR NASSAB RABOR

**MASTER OF SCIENCE
UNIVERSITI PUTRA MALAYSIA
2010**





**MINIMIZATION OF ENTRANCE SURFACE DOSE AND CRITICAL
ORGAN DOSE FOR MEDICAL RADIOGRAPHY USING OPTIMIZATION
PROCEDURES**

By

LOTF ALI MEHDIPOUR NASAB RABOR

**Thesis Submitted to the School of Graduate Studies, Universiti Putra Malaysia,
in Fulfillment of the Requirements for the Degree of Master of Science.**

February 2010



This thesis dedicates to

My dear wife (NAFID),

My parents, my brothers, my sisters,

And

My dear niece (Sahar)

Abstract of thesis presented to the Senate of Universiti Putra Malaysia, in fulfillment of the Requirement for the degree of Master of Science

MINIMIZATION OF ENTRANCE SURFACE DOSE AND CRITICAL ORGAN DOSE FOR MEDICAL RADIOGRAPY USING OPTIMIZE PROCEDURES

By

LOTF ALI MEHDIPOUR NASSAB RABOR

February 2010

Chairman : Professor Dr. Elias Bin Saion, PhD

Faculty : Science

In hospitals, clinics, nursing homes, and medical laboratories, diagnostic radiography is extensively used to view internal structures of a patient and to aid radiologists and doctors diagnose and treat illness and injury. The most common medical radiography is by passing x-rays through the body and captures the image of the patient by means of a sensitized film. This radiographic procedure must feature the system of dose optimization and limitation known as the practice of ALARA (as low as reasonably achievable) recommended by the International Commission on Radiological Protection. The measurements of the entrance surface dose (ESD) and the critical organ dose (COD) in common medical radiography are very important to ensure the dose limit received by the patient in a single radiographic exposure must fulfill the ALARA principle. The most useful way to evaluate ESD is either by indirect measurement on a phantom using an ionization chamber or thermo luminescent dosimeter (TLD) or using calculation based on mathematical model.



In this work, the ESD values measured by the indirect method on RANDO phantom using ionization chamber and by direct method on volunteer patients using TLD. The ESD values of X-ray examinations on chest posterior-anterior (PA) and lateral (lat), skull PA and lat, cervical spine anterior-posterior (AP) and lat, thoracic spine AP and lat, lumbosacral spine AP and lat, abdomen AP, and pelvis AP were found to be 0.15, 0.75, 1.82, 1.16, 0.70, 0.73, 3.6, 5.40, 4.74, 11.7, 4.55, and 3.26 mGy respectively. The IAEA (1996) ESD values are 0.4, 1.5, 5, 3, -, -, 7, 20, 10, 30, 10, and 10 mGy respectively and the NRPB (2000) values are 0.2, 1, 3, 1.6, -, -, 3.5, 10, 6, 14, 6, and 4 mGy respectively. Patients weighting between 65 to 75 kg, from two hospitals in Iran were volunteered to be placed with TLD 100 (LiF) chips during X-ray examinations on chest, skull, cervical, thoracic, lumbar, abdomen, and pelvis to measure the ESD values. For each radiographic procedure 10 patients were selected and in one sachet three TLD chips were used. The measured ESD values on RANDO phantom and patients were compared with the guidance levels provided by the International Atomic Energy Agency (IAEA) and the National Radiological Protection Board (NRPB).

For the indirect measurement of COD values, RANDO phantom and TLDs were used during ESD X-ray measurements. The COD values measured on chest PA for lens, thyroid, and testis were 0.11, 0.12, 0.09 mGy respectively, on chest lat for lens, thyroid and testis were 0.12, 0.14, 0.1 mGy respectively, on skull PA for lens, thyroid and testis were 0.20, 0.19, 0.13 mGy respectively, on skull lat for lens, thyroid and testis were 0.90, 0.20, 0.15 mGy respectively, on cervical spine AP for lens, thyroid and testis were 0.24, 0.67, 0.20 mGy respectively, on cervical spine lat for lens, thyroid and testis were 0.19, 0.62, 0.17 mGy respectively, on thoracic AP

for lens, thyroid and testis were 0.20, 0.68, 0.18 mGy respectively, on thoracic spine lat for lens thyroid and testis were 0.21, 0.25, 0.20 mGy respectively, on lumbosacral spine AP for lens thyroid and testis were 0.24, 0.27, 1.37 mGy respectively, on lumbosacral spine lat for lens thyroid and testis were 0.22, 0.24, 0.40 mGy respectively, on abdomen AP for lens thyroid and testis were 0.23, 0.24, 1.89 mGy respectively, and on pelvis AP for lens thyroid and testis were 0.24, 0.26, 1.23 mGy respectively. In this study the level of stochastic effect of X-rays in diagnostic radiology was lowered in comparison with the guidance levels of IAEA and NRPB. These results enable us to propose the COD values for common general radiography procedures.

Abstrak tesis yang dikemukakan kepada senat Universiti Putra Malaysia sebagai memenuhi keperluan untuk ijazah Master Sains

MENGURANGKAN DOS PERMUKAAN SINAR DATANG DAN DOS ORGAN KRITIKAL UNTUK KEGUNAAN RADIOGRAFI PERUBATAN MENGGUNAKAN KAEDAH PENGOPTIMUMAN

Oleh

LOTF ALI MEHDIPOUR NASSAB RABOR

February 2010

Pengerusi : Profesor Dr. Elias Bin Saion, Ph.D

Fakulti : Sains

Di hospital, klinik, rumah pemulihan dan makmal perubatan, radiograf dianostik kerap digunakan untuk melihat struktur dalaman seorang pesakit dan bertujuan membantu pakar radiologi dan doktor melakukan diagnostik dan merawat penyakit dan kecederaan. Lazimnya radiografi penibutan adalah diperolehi dengan menyinarakan sinar-X menembusi jasad dan merekodkan imej pesakit dengan sekeping filem. Prosedur radiograf perlu menyatakan system pengoptimum dan penghadan dos dikenali sebagai ALARA (serendah mungkin yang boleh dicapai) yang disyorkan oleh Badan Antarabangsa mengenai Perlindungan Sinaran. Pengukuran dos permukaan sinar datang (ESD) dan dos organ kritikal (COD) dalam amalan biasa radiograf perubatan adalah amat penting untuk memastikan had dos yang diterima oleh pesakit dalam pendedahan radiograf tunggal perlu memenuhi prinsip ALARA.lazimny cara paling utama untuk menentukan ESD adalah samada melalui pengukuran terus keatas fentom menggunakan kebuk pengionan atau



dosimeter termopendarcahaya (TLD) atau dengan menggunakan pengiraan berdasarkan model matematik.

Dalam kajian ini, ESD diukur dengan kaedah tak terus menggunakan fantom Rando dan kebuk pengionan dan dengan kaedah terus menggunakan pesakit dan pengesan TLD. Nilai ESD bagi pemeriksaan bagi torkas posterior-anterior (PA) dan lateral (lat), tengkorak PA dan lat, spina servikal anterior-posterior (AP) dan lat, spina torkas AP and lat, spina lumbosakral AP and lat, abdomen AP, and pelvis AP mempunyai nilai masing-masing 0.15, 0.75, 1.82, 1.16, 0.70, 0.73, 3.6, 5.40, 4.74, 11.7, 4.55, dan 3.26 mGy. Nilai ESD bagi IAEA (1996) 0.4, 1.5, 5, 3, -, -, 7, 20, 10, 30, 10, dan 10 mGy manakala nilai untuk NRPB (2000) ialah 0.2, 1, 3, 1.6, -, -, 3.5, 10, 6, 14, 6, dan 4 mGy. Bagi keadah terus, pesakit-pesakit yang beratnya diantara 65 dan 75 kg yang mendapat rawatan daripada dua hospital telah secara sukalera diletakkan cip TLD 100 (LiF) semasa pemeriksaan sinar-X dada, tengkorak, servikal, torkas, lumbar, abdomen, and pelvis to mengukur nilai ESD berkaitan. Untuk setiap kaedah radiograf sebanyak 10 pesakit telah dipilih dan tiga cip TLD diletakkan dalam setiap lokasi pengesan. Nilai-nilai ESD yang diukur dengan menggunakan fantom RANDO dan pesakit telah dibandingkan dengan paras dos yang dicadangkan oleh Agensi Tenaga Atom Antarabangsa (IAEA) dan Badan Perlindungan Sinaran kebangsaan (NRPB).

Untuk mengukur nilai COD secara tak terus, fantom RANDO dan pengesan TLD telah digunakan. Nilai COD yang diukur semasa pemeriksaan sinar-X dada PA untuk kanta, kelenjar tiroid, dan testis adalah masing-masing 0.11, 0.12, 0.09 mGy, peperiksaan dada lat untuk kanta, kelenjar tiroid, dan testis adalah masing-masing 0.12, 0.14, 0.1 mGy, pemeriksaan tengkorak PA untuk kanta, kelenjar tiroid and

testis adalah masing-masing 0.20, 0.19, 0.13 mGy, pemeriksaan tengkorak lat untuk kanta, kelenjar tiroid dan testis adalah masing-masing 0.90, 0.20, 0.15 mGy, pemeriksaan spina servikal AP untuk kanta, kelenjar tiroid dan testis adalah masing-masing 0.24, 0.67, 0.20 mGy, pemeriksaan spina servikal lat untuk kanta, kelenjar tiroid dan testis adalah masing-masing 0.19, 0.62, 0.17 mGy, pemeriksaan spina torkas AP untuk kanta, kelenjar tiroid dan testis adalah masing-masing 0.20, 0.68, 0.18 mGy, pemeriksaan spina torkas lat untuk kanta, kelenjar tiroid dan testis adalah masing-masing 0.21, 0.25, 0.20 mGy, pemeriksaan spina on lumbosacral spine x-ray AP for lens thyroid and testis adalah masing-masing 0.24, 0.27, 1.37 mGy, pemeriksaan spina lumbosakral lat untuk kanta, kelenjar tiroid dan testis adalah masing-masing 0.22, 0.24, 0.40 mGy, pemeriksaan abdomen AP untuk kanta, kelenjar tiroid dan testis adalah masing-masing 0.23, 0.24, 1.89 mGy, dan pemeriksaan pelvis AP untuk kanta, kelenjar tiroid dan testis adalah masing-masing 0.24, 0.26, 1.23 mGy. Dalam kajian ini paras kesan stokastik sinar-X dalam radiology adalah lebih rendah berbanding dengan paras panduan oleh IAEA dan NRPB. Keputusan kajian ini membolehkan kita mencadangkan nilai COD bagi prosedur radiografi am biasa.

ACKNOWLEDGEMENTS

**“Open your heart eye, to see the ghost
You will see any invisible thing
If you split the heart of any particle,
You will see a sun within it,,**

At the end of this step of my graduate period has allowed for a bit of reflection, and the many people who have contributed to both my work, and my life during of this period of time.

First, I would like to express my full thanks and sincere gratitude to my great dear supervisor, Professor Dr Elias Saion for all of guidance, discussions, unlimited assistance and consultations; also I would like to thank my committee members, Mr. Abd. Aziz Mohd Ramli, and Dr Halimah Mohammed Kamari, for their invaluable suggestions, beneficial advices and their endless helps.

I wish to acknowledge my gratitude to all lecturers and staffs in physics department of faculty of science and medical physics group of Malaysian institute of nuclear research technology.

I also wish to acknowledge my gratitude to my dear wife for her encouragements, her emotional supports, and her fortitude. I am grateful to my friend Mohammad Ali Shafahi for his helps; also I would like to express my full thanks and sincere gratitude to my dear parents for their efforts in my life time.

Lotf Ali Mehdipour

2010



I certify that a Thesis Examination Committee has met on 11.2.2010 to conduct the final examination of Lotf Ali Mehdipour nassab rabor on his Master thesis entitled "Minimization of entrance surface dose and critical organ dose for medical radiography using optimization procedures" in accordance with the Universities and University Colleges Act 1971 and the Constitution of the Universiti Putra Malaysia [P.U. (A) 106] 15 March 1998. The Committee recommends that the student be awarded the Master of Science.

Members of the Thesis Examination Committee were as follows:

Zainal Abidin Talib, PhD

Associate Professor
Faculty of Science
Universiti Putra Malaysia
(Chairman)

Abdul Halim Shaari, PhD

Professor
Faculty of Science
Universiti Putra Malaysia
(Internal Examiner)

Azmi Zakaria, PhD

Professor
Faculty of Science
Universiti Putra Malaysia
(Internal Examiner)

Ahmad Termizi Ramli, PhD

Professor
Faculty of Science
Universiti Technology Malaysia
Malaysia
(External Examiner)

BUJANG BIN KIM HUAT, PhD

Professor and Deputy Dean
School of Graduate Studies
Universiti Putra Malaysia

Date:



This thesis was submitted to the senate of Universiti Putra Malaysia and has been accepted as fulfillment of the requirement for the degree of Master of Science. The members of the Supervisory Committee were as follows:

Elias Saion, PhD

Professor
Faculty of science
University Putra Malaysia

Halimah Mohammed Kamari, PhD

Senior Lecturer
Faculty of science
University Putra Malaysia

Abd Aziz Mhd Ramli, PhD

Researcher
Malaysian Nuclear Agency
MINT
Malaysia

HASSANAH MOHD. GHZALI, PhD

Professor and Dean
School of Graduate Studies
University Putra Malaysia
Date: 8 April 2010



DECLARATION

I hereby declare that the thesis is based on my original work except for quotations and citations, which have been duly acknowledged. I also declare that it has not been previously, and is not concurrently, submitted for any other degree at University Putra Malaysia or at any other institution.

(Signature)

LOTF ALI MEHDIPOUR NASSAB RABOR

2010



TABLE OF CONTENTS

| | Page |
|--|-------------|
| ABSTRACT | i |
| ABSTRAK | v |
| ACKNOLEDGMENTS | viii |
| APPROVAL | ix |
| DECLARATION | xi |
| LIST OF TABLES | xv |
| LIST OF FIGURES | xvi |
| LIST OF ABBREVIATIONS | xvii |
| | |
| CHAPTER | |
| 1 INTRODUCTION | 1 |
| 1.1 General introduction | 1 |
| 1.2 Significance of study | 2 |
| 1.3 Problem statement | 3 |
| 1.4 Objective of the study | 4 |
| 1.5 Outline of the thesis | 5 |
| | |
| 2 LITERATURE REVIEW | |
| 2.1 Introduction | 6 |
| 2.2 Biological effects of ionizing radiation | 7 |
| 2.2.1 Stochastic effects | 8 |
| 2.2.2 Deterministic effects | 10 |
| 2.3 Medical Imaging | 12 |
| 2.3.1 Diagnostic x-rays | 12 |
| 2.3.2 X-ray computed tomography | 14 |
| 2.3.3 Nuclear magnetic resonance imaging | 14 |
| 2.3.4 Diagnostic ultrasound | 16 |
| 2.3.5 Nuclear medicine | 16 |
| 2.4 Entrance surface dose | 17 |
| 2.5 Critical organ dose | 19 |
| 2.6 Anthropomorphic phantom | 20 |
| 2.7 Optimization in general radiography | 21 |
| 2.8 ALARA philosophy | 22 |
| | |
| 3 THEORY | |
| 3.1 Medical radiography | 24 |
| 3.2 Ionization effects of radiation | 25 |
| 3.2.1 Electromagnetic radiation | 25 |
| 3.2.2 Production of x-ray | 27 |
| 3.2.3 Bremsstrahlung or braking radiation | 30 |
| 3.2.4 Characteristic radiation | 30 |
| 3.3 Interaction between x-ray photons and matter | 31 |
| 3.3.1 Photoelectric effect | 32 |



| | | | |
|----------|--------|--|----|
| | 3.3.2 | Rayleigh scattering | 35 |
| | 3.3.3 | Compton scattering | 35 |
| 3-4 | | Entrance surface dose (ESD) | 38 |
| | 3.4.1 | Phantom | 39 |
| | 3.4.2 | Ionization chamber | 40 |
| 3-5 | | Factors affecting radiation quality | 41 |
| | 3.5.1 | HVL | 41 |
| | 3.5.2 | kVp | 41 |
| | 3.5.3 | Filtration | 42 |
| | 3.5.4 | Inherent filtration | 42 |
| | 3.5.5 | Added filtration | 42 |
| | 3.5.6 | Collimation | 43 |
| | 3.5.7 | Contrast | 44 |
| 3-6 | | Factors affecting radiographic density | 44 |
| | 3.6.1 | Tube current | 45 |
| | 3.6.2 | Source image receptor distance (SID) | 45 |
| | 3.6.3 | Kilovoltage peak (kVp) | 46 |
| | 3.6.4 | Intensifying screen | 46 |
| | 3.6.5 | Grid | 48 |
| | 3.6.6 | Filtration | 48 |
| | 3.6.7 | Patient factors | 49 |
| | 3.6.8 | Generator type | 50 |
| | 3.6.9 | Beam restriction | 50 |
| | 3.6.10 | Processing | 51 |
| 3-7 | | radiographic contrast | 51 |
| | 3.7.1 | Anode voltage (kVp) | 52 |
| | 3.7.2 | Scattered radiation | 53 |
| | 3.7.3 | Patient factors | 54 |
| | 3.7.4 | Filtration | 55 |
| | 3.7.5 | Intensifying screen | 56 |
| | 3.7.6 | Image receptor (film) | 56 |
| 3.8 | | Critical organ dose (COD) and ALARA philosophy | 58 |
| | 3.8.1 | Shielding | 58 |
| | 3.8.2 | Time | 59 |
| | 3.8.3 | Distance | 59 |
| 4 | | MATERIALS AND METHOD | |
| 4.1 | | Materials | 61 |
| | 4.1.1 | Phantom: Rando Alderson Phantom | 63 |
| | 4.1.2 | Radiography X-ray unit | 62 |
| | 4.1.3 | Ionization chamber | 65 |
| | 4.1.4 | Thermo luminescent dosimeter | 66 |
| 4.2 | | Instrumentation | 66 |
| | 4.2.1 | Humidity meter | 67 |
| | 4.2.2 | Barometer and thermometer | 67 |
| | 4.2.3 | Processing system | 68 |
| | 4.2.4 | Film-screen combination | 68 |
| | 4.2.5 | TLD reader system | 70 |
| | 4.2.6 | Annealing system | 70 |

| | | |
|----------|--|-----|
| 4.3 | Methods | 72 |
| 4.3.1 | Measurement of ESD on phantom | 75 |
| 4.3.2 | Calibration of ionization chamber | 74 |
| 4.3.3 | Geometry used for calculation of ESD | 76 |
| 4.3.4 | Dose measurement in common Radiography | 77 |
| 4.3.5 | Quality control of radiographs | 82 |
| 4.3.6 | Measurement of critical organ dose by TLD | 86 |
| 4.3.7 | Calibration method of TLD | 87 |
| 4.3.8 | Geometry used for calculation of critical organ dose | 88 |
| 4.3.9 | Critical organ dose measurement | 90 |
| 4.3.10 | Measurement of ESD on patient | 90 |
| 5 | RESULTS AND DISCUSSION | |
| 5-1 | Radiation quality measurement | 92 |
| 5.1.1 | HVL measurement | 92 |
| 5.1.2 | Measurement of effect of kVp on ESD | 99 |
| 5-2 | Radiation quantity measurement | 107 |
| 5.2.1 | Measurement of effect of mAs on ESD | 107 |
| 5.2.2 | Measurement of effect of SID on ESD | 111 |
| 5-3 | Measurement of Entrance surface dose (ESD) | 117 |
| 5.3.1 | Comparative ESD measurement on patient | 120 |
| 5-4 | Measurement of critical organ dose (COD) | 123 |
| 5.4.1 | Critical organ dose (Lens) | 127 |
| 5.4.2 | Critical organ dose (Thyroid) | 128 |
| 5.4.3 | Critical organ dose (Testis) | 129 |
| 6 | CONCLUSION AND FUTURE WORK | |
| 6-1 | Conclusion | 131 |
| 6-2 | Future work and recommendations | 132 |
| 6.2.1 | Speed of Film | 133 |
| 6.2.2 | Critical Organ Dose (COD) | 133 |
| | REFERENCES | 134 |
| | BIODATA OF STUDENT | 139 |

LIST OF TABLES

| Table | | Page |
|-------|--|------|
| 4.1 | calibration coefficient of ionization chamber in different kVp | 76 |
| 4.2 | evaluation criteria for some common medical radiography | 83 |
| 5.1 | suggested radiation factors for some common medical radiography | 116 |
| 5.2 | suggested entrance surface dose (ESD) values in mGy in present study on phantom and comparison with the recommended values of IAEA and NRPB. | 119 |
| 5.3 | comparison of ESD (mGy) measurements in patients, Rando phantom and values recommended from IAEA (1996) and NRPB (2000). | 122 |
| 5.4 | The critical organ (lens, thyroid and testis) doses (mGy) in general radiography measured by TLD on Rando phantom. | 126 |



LIST OF FIGURES

| Page | | Figure |
|------|--|--------|
| 3.1 | Electromagnetic spectrum | 26 |
| 3.2 | Modern x-ray tube | 29 |
| 3.3 | Diagram of photoelectric effect | 34 |
| 3.4 | Photo electric effect | 35 |
| 3.5 | Schematic Diagram of Compton scattering | 36 |
| 3.6 | Schematic kinematic of Compton scattering | 37 |
| 4.1 | Rando phantom of man used in this work | 62 |
| 4.2 | Rando phantom is made up from thirty-five 2.5 cm thick transverse slices | 62 |
| 4.3 | Radiography X-ray machine | 64 |
| 4.4 | Control panel of x-ray machine | 64 |
| 4.5 | Ionization chamber (Radcal 9010) used in this work | 65 |
| 4.6 | Vacuum tweezers | 66 |
| 4.7 | Humidity meter | 67 |
| 4.8 | Barometer and thermometer | 68 |
| 4.9 | Image processor machine | 69 |
| 4.10 | Radiographic cassette | 69 |
| 4.11 | TLD readers Harshow 3500 | 71 |
| 4.12 | Annealing system | 71 |
| 4.13 | Annealing process of TLD | 72 |
| 4.14 | Geometry used for measurement of ESD | 76 |
| 4.12 | Annealing system | 68 |
| 4.13 | Annealing process of TLDs | 69 |

| | | |
|------|--|-----|
| 4.14 | Geometry used for measurement of ESD | 77 |
| 4.15 | Geometry used for measurement of critical organ dose | 89 |
| 4.16 | Critical organ dose measurement by TLD | 89 |
| 5.1 | Out put with, Al – thickness, 55 kVp | 93 |
| 5.2 | Out put with, Al – thickness, 65 kVp | 94 |
| 5.3 | Out put with, Al – thickness, 70 kVp | 94 |
| 5.4 | Out put with, Al – thickness, 80 kVp | 95 |
| 5.5 | Out put with, Al – thickness, 90 kVp | 95 |
| 5.6 | Out put with, Al – thickness, 100 kVp | 96 |
| 5.7 | Out put with, Al – thickness, 110 kVp | 96 |
| 5.8 | X-ray Out put with, Al – thickness, for 55 kVp to 110 kVp | 98 |
| 5.9 | The linear Relation between HVL and applied anode voltage kVp | 99 |
| 5.10 | The entrance surface dose (ESD) (mGy) versus anode voltage (kV) in AP projection of abdomen X-rays. | 100 |
| 5.11 | Radiographic image taken at 46 mAs and 70 kV; SID 100 cm, field size 34×42 cm ² , film type green, film speed 400 | 102 |
| 5.12 | Radiographic image taken at 6 mAs and 80 kV; SID 180 cm, field size 34 × 36 cm ² , film type green, film speed 400 (chest x-ray (PA | 103 |
| 5.13 | The entrance surface dose (ESD) (mGy) versus anode voltage (kV) in PA projection of chest X-rays. | 104 |
| 5.14 | The location of heart and lungs in chest region of human body | 105 |
| 5.15 | Comparative evaluations in ESD (mGy) versus voltage kVp between AP abdomen X-rays and PA chest X-rays | 106 |
| 5.16 | Relation between ESD (mGy) and mAs in abdomen (AP). | 108 |
| 5.17 | Relation between ESD (mGy) and mAs in Chest (PA). | 109 |

| | | |
|------|---|-----|
| 5.18 | Comparison between Chest (PA) and Abdomen (AP) examinations Showing the radiation quality used. | 110 |
| 5.19 | Radiographic image of Skull (PA) | 111 |
| 5.20 | Relation between ESD (mGy) and SID (cm). | 112 |
| 5.21 | Relation between SID (Cm) and ESD (mGy). | 114 |
| 5.22 | Comparison between chest (PA) and skull (PA) examinations | 115 |
| 5.23 | Measured ESD values of various radiograph evaluations compared. With recommended values from IAEA (1996) and NRPB (2000). | 117 |
| 5.24 | Measured ESD values of various radiograph evaluations on patient compared with the ESD values that are measured on RANDO phantom in this study. | 123 |
| 5.25 | Measured lens COD (mGy) in 12 procedures. | 127 |
| 5.26 | Measured thyroid COD (mGy) in 12 procedures | 129 |
| 5.27 | Measured testis COD (mGy) in 12 procedures | 130 |

LIST OF ABBREVIATIONS

| | |
|-------|--|
| ALARA | As Low As Reasonably Achievable |
| AP | Anterior - Posterior |
| BG | Back ground radiation |
| CF | Calibration Factor |
| COD | Critical Organ Dose |
| Cm | Centimeter |
| CPS | Cycle per second |
| DNA | Deoxyribonucleic Acid |
| DRL | Diagnostic Reference Level |
| ESD | Entrance Surface Dose |
| HVL | Half Value Layer |
| HZ | Hertz |
| IAEA | International Atomic Energy Agency |
| ICRP | International commission of Radiation Protection |
| KERMA | Kinetic Energy Released in Material |
| Kev | Kiloelectron volt |
| Kpa | Kilopascal |
| kVp | Kilovolt peak |
| Lat | Lateral |
| LiF | Lithium Florid |
| mAs | milliampere second |
| mbar | millibar |
| MDD | minimum Detectable Dose |



| | |
|--------------|---|
| mGy | milligray |
| MINT | Malaysian Institute of Nuclear research Technology |
| MM | millimeter |
| mR | milliroentgen |
| NC | nanocoulomb |
| PA | Posterior - Anterior |
| Rad | Radiation Absorbed Dose |
| RPM | Round per minute |
| SID | Source Image Distance |
| SSDL | Secondary Standard Dose Laboratory |
| TLD | Thermo Luminescent Dosimeter |
| UK | United Kingdom |
| USA | United State of America |
| μ Gy | microgray |
| $^{\circ}$ C | degree centigrade |
| $^{\circ}$ F | degree fahrenheit |

CHAPTER 1

INTRODUCTION

1.1 General introduction

X-ray was discovered in the 1895 by Wilhelm Conrad Roentgen by accident. He spent several weeks working in his laboratory to investigate the properties of X-rays and noticed that when he placed his hand between his energized tube and the barium platinocyanide-coated paper, he could see the bones of his hand glow on the paper with the fluoroscopic image moving as he moved his hand. He also produced a static radiograph of his wife's hand using this new radiation (Alexi A, 1995). This was the first radiograph of the world and has become an important step in physics and medicine. He proceeded to study them so thoroughly that within a very short period of time, he had identified the properties of X-rays that we recognized today.

In medical diagnostic radiology, X-ray creates images help diagnose the patient's medical conditions. The quality of the images produced depends partly on how much radiation is used, which is under the direct control of radiographers and radiologists. To produce a quality X-ray image, the amount of radiation transverse the patient depends on various factors such as type of examination, patient's physical characteristics, and condition of pathology to be imaged and several factors related to

instrumentation. One of the major goals of medical radiography is to minimize radiation dose to the patient without compromising the image quality needed to produce an accurate diagnosis (Seeram, and Patrick., 2006). This is the subject of the present study.

1.2 Significance of study

The basic principles of radiological protection as suggested by the International Commission of Radiological Protection (ICRP) are justification of the practice and optimization of protection. Justification is the first step in radiological protection, which has been accepted that, diagnostic exposure is justifiable only when there is a valid clinical indication, no matter how good the imaging performance may be, that every examination must result in a net benefit to the patient (IAEA, 2004). In the area of optimization in diagnostic radiology there is considerable scope for reducing dose without loss of diagnostic information, but the extent to which the measures available are used varies widely. The optimization of protection in diagnostic radiology does not necessarily mean the reduction of doses to the patient. It is paramount that the image obtained contains the diagnostic information as intended (IAEA, 2004).

The need to optimize the protection of patients results from the fact that medical exposures are by far the largest source of exposition of artificial origin. Diagnostic