



UNIVERSITI PUTRA MALAYSIA

**POLYMERIC MEMBRANE SENSORS FOR DETECTION OF
TITANIUM (III) IONS BASED ON TRIPODAL- OR
CALIXARENE-LIGAND AND DETECTION OF
CHROMIUM (VI) USING METALLO-SALEN
IONOPHORES**

MAJID REZAYI

FS 2011 15



**POLYMERIC MEMBRANE SENSORS FOR DETECTION OF
TITANIUM (III) IONS BASED ON TRIPODAL- OR
CALIXARENE-LIGAND AND DETECTION OF
CHROMIUM (VI) USING METALLO-SALEN
IONOPHORES**

By

MAJID REZAYI

**Thesis submitted to the School of Graduate Studies, Universiti Putra
Malaysia, in Fulfilment of the Requirements for the Degree of
Doctor of Philosophy**

May 2011



Special dedication to my beloved parents



Abstract of thesis presented to the Senate of Universiti Putra Malaysia in fulfilment of the requirement for the degree of Doctor of Philosophy

POLYMERIC MEMBRANE SENSORS FOR DETECTION OF TITANIUM (III) IONS BASED ON TRIPODAL- OR CALIXARENE-LIGAND AND DETECTION OF CHROMIUM (VI) USING METALLO-SALEN IONOPHORES

By

MAJID REZAYI

May 2011

Chair: Professor Anuar Kassim, PhD

Faculty: Science

Currently, despite of many developments in the field of ion selective electrodes (ISEs), it continues to evolve. The introduction of low detection limit ISEs may open a new opportunity to determine trace target ions. Improvements of detection limits, selectivities, understanding the response mechanism and developing new membrane materials are constantly being reported in the literatures. Moreover, the development of ISEs used in hospital and industrial setting as well as in clinical, environmental and physiological research has been opened new horizons in front of us. This research is focused on the fabrication of new PVC-membrane sensors based on tris(2pyridyl) methylamine, tpm and c-methylcalix[4]resorcinarene, CMCR as an cationic ionophores for detection of titanium (III) cation, and also N,N'Bis(salicylidene) ethylenediamino cobalt(II), Co(SALEN) as an anionic ionophore for determining the chromate (II) anion.

The complex reactions of Co(SALEN) with chromate (II) anion, tpm and CMCR with titanium (III) cation in water (H₂O), acetonitrile (AN) and their binary mixture solutions at different temperatures by using the conductometric method, are investigated. The characterization and evaluation of materials are described. In addition, a variety of analytical methods, including UV-Vis spectroscopy, FT-IR spectroscopy, Scanning Electron Microscopy (SEM), are used to study these processes. Later on, the application and validation of proposed ISEs with the potentiometric titration, atomic absorption spectrometry (AAs) and inductively coupled plasma atomic emission spectrometry (ICP-AES) are studied. Based on conductometric measurement results, the stoichiometry of complex formation for all the cases of ion-ionophore is 1:1. Furthermore, the average of stability constant ($\log K_f$) obtained for tpm-Ti(OH)(OH₂)₅²⁺, CMCR-Ti(OH)(OH₂)₅²⁺ and Co(SALEN)-CrO₄²⁻ complexes at 25°C are 2.70, 3.13 and 3.02, respectively. Therefore, the mentioned ionophores can be used as sensing elements to fabricate ISE membranes for determination of titanium and chromate ions. These electroactive composite materials resulted in three new ionophore types which one based on selective ion-sensing membrane electrodes. They were fabricated for the determination of Ti (III) and CrO₄ (II) ions in solutions. The membrane sensors showed fast, stable and Nernstian response for the cation of titanium (III) and anion of chromate (II) over the concentration range from 1.0×10⁻⁶ to 1.0×10⁻² M and 1.0×10⁻⁶ to 1.0×10⁻¹ M at 25°C, over the pH range from 1 to 2.5 and 7 to 10 respectively. For titanium (III) cation, based on tpm and CMCR ionophores, the Nernstian slopes, detection limits and response times were 29.17±0.24 and 30.38±0.15 mV/decade of activity, 7.9×10⁻⁷ and 8.9×10⁻⁷ M, 20 and 15 s, respectively. For chromate (II)

anion, using Co(SALEN) ionophore, the Nernstian slope, detection limit and response time was -28.33 ± 0.10 mV/decade of activity, 7.9×10^{-7} M and <10 s. The direct determination of 4 to 39 $\mu\text{g/ml}$ of titanium (III) standard solution and 2 to 48.5 $\mu\text{g/ml}$ of chromate (II) standard solution showed an average recovery of 94.60, 94.70 and 96.03 % and a mean relative standard deviation of 1.8, 2.2 and 1.6% at 100.0 $\mu\text{g/ml}$ for tpm, CMCR and Co(SALEN) sensors, respectively. Finally, the utilizing of electrodes as the end point indicators for potentiometric titration with EDTA and $\text{Pb}(\text{NO}_3)_2$ solutions for titanium (III) and chromate (II) sensors were successfully carried out respectively.

Abstrak tesis yang dikemukakan kepada Senat Universiti Putra Malaysia
sebagai memenuhi keperluan untuk ijazah Doktor Falsafah

**SENSOR MEMBRAN POLIMERIK UNTUK PENGESANAN ION TITANIUM
(III) BERDASARKAN TRIPODAL- ATAU KALIKSARENA- LIGAN DAN
PENGESANAN KROMIUM (VI) MENGGUNAKAN IONOFOR
METALLO-SALEN**

Oleh

MAJID REZAYI

Mei 2011

Pengerusi: Profesor Anuar Kassim, PhD

Fakulti: Sains

Pada masa kini, perkembangan dalam bidang elektrod pemilih-ion (ISE) masih berterusan. Pengenalan ISEs dengan had pengesanan rendah ini membuka peluang baru untuk menentukan ion sasaran surih. Peningkatan had pengesanan, selektiviti, pemahaman mekanisma tindak balas dan perkembangan bahan membran baru secara berterusan telah dilaporkan di dalam banyak penerbitan literatur. Selain itu, pembangunan ISE untuk digunakan dalam hospital dan industri seperti dalam kajian klinikal, persekitaran dan fisiologi telah membuka era baru dalam bidang ini. Kajian ini difokuskan kepada pembuatan/fabrikasi deria (sensor)PVC-membran baru berdasarkan pada bahan tris(2piridil) metilamina, tpm dan c-metilcalix[4]resorcinarena, CMCR sebagai ionofor kationik untuk mengesan kation titanium (III), dan juga

N,N'Bis(salicilidena) etilenadiamino kobalt (II), Co(SALEN) sebagai ionofor anionik untuk menentukan anion kromat (II). Kompleks tindakbalas Co(SALEN) dengan anion kromat (II), tpm dan CMCR dengan kation titanium (III) dalam air (H₂O), asetonitril (AN) dan larutan campuran binari kepada suhu yang berbeza dengan menggunakan kaedah konduktometri telah dikaji. Pencirian dan penilaian bahan dibincangkan. Selain itu, pelbagai kaedah analisis, termasuk serapan UV-Vis, spektroskopi FT-IR, Imbasan Mikroskop Elektron (SEM), digunakan dalam kajian proses ini. Seterusnya, aplikasi dan pengesahan untuk ISE yang dicadangkan, dikaji dengan pentitratan potentiometri, serapan atom (AA) dan induktif kupel plasma (ICP) spektrometri. Berdasarkan hasil pengukuran konduktometri, stoikiometri pembentukan kompleks untuk semua kes ion-ionofor adalah 1:1. Selain itu, pemalar kestabilan ($\log K_f$) purata yang diperolehi, untuk tpm-Ti(OH)(OH₂)₅²⁺, CMCR-Ti(OH)(OH₂)₅²⁺ dan Co(SALEN)-CrO₄²⁻ kompleks pada suhu 25°C adalah 2.70, 3.13 dan 3.02, masing-masing. Oleh kerana itu, ionofor tersebut boleh digunakan sebagai penderiaan unsur untuk membuat membran ISE bagi menentukan ion titanium dan kromat. Bahan komposit elektroaktif ini menghasilkan tiga jenis ionofor baru yang berdasarkan kepada membran pemilih ion-deria elektrod. Ionofor ini disediakan untuk penentuan Ti (III) dan CrO₄ (II) ion dalam larutan. Sensor membran menunjukkan tindakbalas cepat, stabil dan rangsangan Nernstian untuk kation titanium (III) dan anion kromat (II) dalam lingkungan kepekatan dari 1.0×10⁻⁶ hingga 1.0×10⁻² M dan 1.0×10⁻⁶ hingga 1.0×10⁻¹ M pada 25°C, pada lingkungan pH 1 hingga 2.5 dan 7 hingga 10 masing-masing. Untuk kation titanium (III), berdasarkan ionofor tpm dan CMCR, kecerunan Nernst, had pengesanan dan masa tindakbalas adalah 29.17±0.24 dan 30.38±0.15 mV/dekad aktiviti, 7.9×10⁻

⁷ dan 8.9×10^{-7} M, 20 dan 15 s, masing-masing. Untuk anion kromat (II), yang menggunakan Co(SALEN) ionofor, kecerunan Nernst, had pengesanan dan masa tindakbalas adalah -28.33 ± 0.10 mV /dekad aktiviti, 7.9×10^{-7} M dan <10 s. Penentuan langsung bagi 4 hingga 39 $\mu\text{g/ml}$ titanium (III) larutan piawai dan 2 hingga 48.5 $\mu\text{g/ml}$ kromat (II) larutan piawai menunjukkan perolehan semula purata 94.60, 94.70 dan 96.03% dengan sisihan piawai purata relatif 1.8, 2.2 dan 1.6% pada 100.0 $\mu\text{g/ml}$ untuk tpm, CMCR dan Co(SALEN) sensor, masing-masing. Akhirnya, penggunaan elektrod sebagai penunjuk untuk titratan potensiometri dengan larutan EDTA dan $\text{Pb}(\text{NO}_3)_2$ menggunakan elektrod sensor titanium (III) dan kromat (II) masing-masing telah berjaya dilakukan.

ACKNOWLEDGEMENTS

In the name of Allah, the most merciful, the most compassionate:

My God! Let me out of the darkness of illusion, and let me be honourable by the light of understanding. My God, open to us the doors of thy compassion and uncover to us the treasures of the knowledge by thy compassion, the most compassionate of the compassionate ones.

This thesis is not an individual product; it is the result of much support for others.

First, I am ever grateful to the Almighty for being my guiding light in the preparation and writing of this thesis and throughout my life.

I would like to express my appreciation to my supervisory committee, Prof. Dr. Anuar Kassim, Prof. Dr. Lee Yook Heng, Assoc. Prof. Dr. Tan Wee Tee and Assoc. Prof. Dr. Nor Azah Yusof. I have gained a lot of experience and knowledge from your kind guidance.

I am also very grateful to my parents, who were my first teachers. You always have a piece in everything I do, and most often the best. I do not believe that this project could have been done without your phone calls, love, encouragement, and support, which sustained me each and every week.

My good classmates, labmates, and friends, you have filled my life every day and inspired me. I am a better person for having shared my life with you.

My dear housemates thank you for so much happiness during these four years of living together. I spent unforgettable moments in my life with you. Your kind support, your precious friendship, your help, and the good times that we enjoyed together will be valuable memories forever. I am a lucky person for meeting all of you!

For any others that I forgot to mention, I am humbly indebted!

I certify that a Thesis Examination Committee has met on 03 MAY 2011 to conduct the final examination of Majid Rezayi on his thesis entitled “Polymeric membrane sensors for detection of titanium (III) ions based on tripodal- or calixarene-ligand and detection of chromium (VI) using metallo-salen ionophores” in accordance with the Universities and University Colleges Act 1971 and the Constitution of the Universiti Putra Malaysia [P.U.(A) 106] 15 March 1998. The Committee recommends that the student be awarded the Doctor of Philosophy.

Members of the Thesis Examination Committee were as follows:

Sidik Silong, PhD

Associate Professor
Faculty of Science
Universiti Putra Malaysia
(Chairman)

Md. Jelas Haron, PhD

Professor
Faculty of Science
Universiti Putra Malaysia
(Internal Examiner)

Mohamad Zaki Ab. Rahman, PhD

Associate Professor
Faculty of Science
Universiti Putra Malaysia
(Internal Examiner)

Mei Xian Li, PhD

Associate Professor
Faculty of Science
Peking University
(External Examiner)

BUJANG KIM HUAT, PhD

Professor and Deputy Dean
School of Graduate Studies
Universiti Putra Malaysia

Date:

This thesis was submitted to the Senate of Universiti Putra Malaysia and has been accepted as fulfilment of the requirement for the degree of Doctor of Philosophy. The members of the Supervisory Committee were as follows:

Anuar Kassim, PhD

Professor
Faculty of Science
Universiti Putra Malaysia
(Chairman)

Lee Yook Heng, PhD

Professor
Faculty of Science and Technology
Universiti Kebangsaan Malaysia
(Member)

Tan Wee Tee, PhD

Associate Professor
Faculty of Science
Universiti Putra Malaysia
(Member)

Nor Azah Yusof, PhD

Associate Professor
Faculty of Science
Universiti Putra Malaysia
(Member)

HASANAH MOHD GHAZALI, PhD

Professor and Dean
School of Graduate Studies
Universiti Putra Malaysia

Date:

DECLARATION

I declare that the thesis is my original work except for quotations and citations which have been duly acknowledged. I also declare that it has not been previously, and is not concurrently, submitted for any other degree at Universiti Putra Malaysia or at any other institutions.

MAJID REZAYI

Date: 03 May 2011

TABLE OF CONTENTS

	Page
ABSTRACT	iii
ABSTRAK	vi
ACKNOWLEDGEMENTS	ix
APPROVAL	xi
DECLARATION	xiii
LIST OF TABLES	xviii
LIST OF FIGURES	xx
LIST OF ABBREVIATIONS	xxvi
CHAPTER	
1 INTRODUCTION	1
1.1 General Introduction	1
1.2 Background of the Study	4
1.3 Research Problem	6
1.4 Research Objectives	8
1.5 Outline of Thesis	9
2 LITERATURE REVIEW	11
2.1 History of ISEs	11
2.1.1 Titanium Selective Electrode	14
2.1.2 Chromate Selective Electrode	15
2.2 Potentiometric Sensors	18
2.2.1 Ion Selective Electrodes (ISEs)	18
2.2.2 Coated Wire Electrodes (CWEs)	19
2.2.3 Field Effect Transistors (FETs)	20
2.2.4 All Solid State Ion Selective Electrodes (ASSISEs)	21
2.3 Response mechanism of ISEs	22
2.3.1 Dynamic Model	24
2.3.2 Phase Boundary Potential Model	24
2.3.3 The Nernst Equation	27
2.3.4 The Nicolskii-Eisenman Equation	28
2.4 Characterization of ISE	29
2.4.1 Selectivity	29
2.4.2 Detection Limit of ISEs	34
2.4.3 Measuring Range of ISEs	36
2.4.4 Response time of ISEs	36
2.4.5 Life time of ISEs	37
2.4.6 Reproducibility of ISEs	38
2.5 The composition of ISE membrane	38
2.5.1 The Polymeric Matrix	38
2.5.2 The Ionophore (Membrane-Active Recognition)	39
2.5.3 Plasticizer	49
2.5.4 Lipophilic Ionic Additives	50



3	THERMODYNAMIC STUDIES OF COMPLEX FORMATION BETWEEN tpm, CMCR AND Co(SALEN) IONOPHORES WITH TITANIUM (III) AND CHROMATE (II) IONS IN AN-H₂O BINARY SOLUTIONS BY THE CONDUCTOMETRIC METHOD	53
3.1	Introduction	53
3.1.1	Molar and Equivalent Conductivity	55
3.1.2	Cell Constant	57
3.1.3	Solvation and Gutman Donor Number	57
3.1.4	Equilibrium Constant K_f of the Complex Formation	59
3.1.5	Thermodynamic Functions of the Complex Formation	60
3.1.6	GENPLOT Computer Program	62
3.2	Materials and methods	62
3.2.1	Experimental	62
3.2.2	Apparatus	63
3.2.3	Methodology	63
3.3	Results and Discussion	64
3.3.1	Study of Molar Conductivity versus the Molar Ratio of Ligand-to-Ion at Different Temperatures	65
3.3.2	Study of Complex Formation at Different Temperatures	74
3.3.3	Calculation of Thermodynamic Parameters	78
3.3.4	Study of Stability Constant of Complexes Versus Composition of Binary Mixed Solution	80
3.3.5	Van't Hoff plots	82
3.4	Conclusion	84
4	CATION SELECTIVE ELECTRODE FOR TITANIUM (III) BASED ON TPM AND CMCR IONOPHORES	86
4.1	Introduction	86
4.1.1	Titanium Components Chemistry	87
4.1.2	Preparation Methods of Liquid Membrane ISEs	92
4.2	Materials and Methods	93
4.2.1	Experimental	93
4.2.2	Apparatus	94
4.2.3	Methodology	95
4.3	Results and Discussion	102
4.3.1	Working Concentration Range and Slope	102
4.3.2	Effect of Membrane Composition	103
4.3.3	Plasticizer Effect	107
4.3.4	Effect of Anionic Additives	109
4.3.5	Calibration Curve and Statistical Data	110
4.3.6	Repeatability and Reproducibility	111
4.3.7	Effect of Internal Solution	112
4.3.8	Potential Responses of Fabricated Electrodes for Other Cations	114
4.3.9	Effect of pH	119
4.3.10	Dynamic and Static Response times	120
4.3.11	Life time	123



4.3.12	UV-Vis Studies	125
4.3.13	FT-IR Studies	127
4.3.14	X-ray Studies	129
4.3.15	SEM Studies	130
4.3.16	Potentiometric Selectivity of the Sensors	133
4.3.17	Analytical Application	136
4.4	Conclusion	139
5	ANION SELECTIVE ELECTRODE FOR CHROMATE (II) BASED ON CO(SALEN) IONOPHORE	141
5.1	Introduction	141
5.1.1	Chromium Components Chemistry	144
5.1.2	Interaction of Co(SALEN) Ligand with Chromate Anion	146
5.2	Material and Methods	148
5.2.1	Experimental	148
5.2.2	Apparatus	149
5.2.3	Methodology	149
5.3	Results and Discussion	152
5.3.1	Optimization of Membranes Composition	152
5.3.2	Plasticizer Effect	155
5.3.3	Working Concentration Range and Slope	156
5.3.4	Effect of Internal Solution	156
5.3.5	Potential Responses of Fabricated Electrode for Other Anions	157
5.3.6	pH Effect	160
5.3.7	Dynamic and Static Response times	160
5.3.8	Lifetime of the Electrode	162
5.3.9	Selectivity of the Electrode	163
5.3.10	UV-Vis study	165
5.3.11	FT-IR study	166
5.3.12	SEM study	167
5.3.13	Analytical Application	169
5.3.14	Comparison of Chromate-Selective Electrode in This Work with Other Reported Chromate- Selective Electrodes	170
5.4	Conclusion	172
6	SUMMARY, GENERAL CONCLUSION AND RECOMMENDATIONS FOR FUTURE RESEARCH	173
6.1	Summary	173
6.2	General Conclusion	175
6.3	Recommendations for Future Research	176
	REFERENCES	179
	APPENDICES	199
	BIODATA OF STUDENT	207



