



UNIVERSITI PUTRA MALAYSIA

**DETERMINATION OF OIL PALM LEAF NITROGEN CONCENTRATION
BY REFLECTANCE SPECTROSCOPY**

AHMAD SARWANI HAJI ADNI

FP 2010 35

**DETERMINATION OF OIL PALM LEAF NITROGEN CONCENTRATION BY
REFLECTANCE SPECTROSCOPY**



By

AHMAD SARWANI HAJI ADNI

**Thesis Submitted in Fulfilment of the Requirements for the Degree of
Master of Science**

January 2010

DEDICATION

This thesis is dedicated to all staff and top management for their support



Abstract of the thesis presented to the Senate of the Universiti Putra Malaysia in fulfillment of the requirement for the degree of Master of Science

DETERMINATION OF OIL PALM LEAF NITROGEN CONCENTRATION BY REFLECTANCE SPECTROSCOPY

By

AHMAD SARWANI HAJI ADNI

January 2010

Chairman: Dr. Anuar Abdul Rahim, PhD

Faculty : Agriculture

Nitrogen (N) presents in plant materials usually in the form of protein and its analysis in the laboratory usually involve digestion process. In the digestion process, protein nitrogen is converted to the ammonium form by heating with concentrated sulphuric acid in the presence of a catalyst mixture of sodium sulphate and selenium. The method for determining the final ammonium may be either by distillation, as in the classical Kjeldahl method, or by an instrumental analysis. In recent years, a new method has been developed to measure N content in leaves using a fiber optic spectrometer. This method allows N content in leaves to be determined in situ and it can provide real time data of N content in leaves. Compared to the traditional destructive methods, this method will reduce the cost of analyzing the plant nutrient contents by minimizing chemical cost, time and labor requirements. Furthermore, since this technique does not involve the use of chemicals, there will be no generation of chemical waste. An example of this new method is the miniature fiber optic spectrometer with specially designed leaf probe, the Ocean Optic HR2000 High Resolution Miniature Fiber Optic Spectrometer. However, to get

accurate data from this probe, a relationship between the data obtained from the sensor must be correlated with actual foliar nutrient levels measured in the laboratory. Very few studies have been conducted to determine the relationship between the N data obtained using the leaf probe and the N level measured in the laboratory.

The objective of this study, therefore, was to determine the relationship between the reflectance measurements using the Ocean Optic HR2000 High Resolution Miniature Fiber Optic Spectrometer with N concentration determined using the traditional wet-chemistry procedure (Kjeldahl digestion technique) from fresh oil palm leaf of different fertilization treatments and from fresh oil palm leaf of different ages. Significant relationships were found between reflectance of fresh oil palm leaves with N from Kjeldahl method. The N concentration in the leaf samples of oil palm from different ages analyzed by the Kjeldahl technique ranged from 1.89 % to 3.24 %. When the oil palm leaves from each field were analyzed separately, the highest relationships were measured in the red region of the spectrum between 699.98 to 700.87 nm. The use of pooled data to obtain the regression equations resulted in precision greater than that achieved with the single-field data. The best regression line in predicting leaf N concentration from the measured reflectance was obtained using the combined (pooled) data and the equation was $\text{Leaf N} = 2.7935 - 0.0014 \text{ Reflectance}$. The difference between the mean of actual N concentration and the mean of predicted N concentration from pooled data was only 0.031%, which was not statistically significant ($P \geq 0.05$, $df=266$) according to the performed t-test. The results indicate that the reflectance spectroscopy method using the Ocean Optic HR2000 High Resolution Miniature Fiber Optic Spectrometer was reliable for the determination of N concentration in fresh oil palm leaves.

Abstrak tesis yang dikemukakan kepada Senat Universiti Putra Malaysia sebagai memenuhi keperluan untuk penganugerahan ijazah Master Sains

**PENENTUAN KANDUNGAN NITROGEN DALAM DAUN KELAPA SAWIT
MELALUI SPEKTROSKOPI KEPANTULAN**

Oleh

AHMAD SARWANI HAJI ADNI

January 2010

Pengerusi: Dr. Anuar Abdul Rahim, PhD

Fakulti: Pertanian

Nitrogen (N) biasanya hadir di dalam daun sebagai protin dan analisa N di makmal melibatkan proses penghancuran. Dalam proses ini, protin N ditukarkan kepada bentuk ammonium secara memanaskannya dengan asid hidrogensulfat dengan kehadiran campuran pemangkin terdiri dari garam sodium sulfat dan selenium. Penentuan kandungan ammonium akhir dilakukan samada secara kaedah penyulingan, seperti dalam kaedah tradisi Kjeldahl, atau melalui kaedah analisa peralatan. Beberapa tahun kebelakangan ini, satu kaedah baru menentukan kandungan N menggunakan alat spectrometer gentian optic telah dibangunkan. Melalui kaedah ini, kandungan N dapat dianalisa secara terus di lapangan. Berbanding kaedah lama yang merosakkan, kaedah ini dapat mengurangkan kos analisa kandungan nutrisi kelapa sawit dengan cara mengurangkan kos bahan kimia, masa dan keperluan tenaga kerja. Selain dari itu, oleh kerana teknik ini tidak memerlukan penggunaan bahan kimia, tiada sisa buangan kimia yang dihasilkan. Satu contoh kaedah baru ini ialah dengan menggunakan spectrometer

mini yang direka dengan alat deria daun, HR2000 Mini Spektrometer Berresolusi Tinggi model Ocean Optic. Walau bagaimanapun, untuk mendapatkan data yang tepat, satu bentuk hubungan di antara data dari alat deria dan kandungan nutrisi daun dari analisa makmal perlu dibena. Hanya beberapa penyelidikan telah dijalankan untuk menentukan hubungan tersebut. Oleh yang demikian, kajian ini bertujuan mendapatkan hubungan di antara data kepantulan melauai Spektrometer Mini HR 2000 Gentian Optik Berresolusi Tinggi model Ocean Optic dan peratusan kandungan N melalui analisa kimia di dalam makmal (teknik Kjeldahl) pokok kelapa sawit segar berbeza pembajaan dan daripada pokok pelbagai usia.

Hasil ujian didapati, hubungan yang signifikan telah diperolehi di antara data pantulan daun kelapa sawit segar dengan kandungan N melalui kaedah Kjeldahl dengan kandungan N di antara 1.89 % to 3.24 %. Apabila daun kelapa sawit dari lapangan berbeza di analisa secara berasingan, hubungan tertinggi diperolehi dalam kawasan spectrum merah di antara 699.98 dan 700.87 nm. Penggunaan persamaan regresi hasil data terkumpul menghasilkan keputusan yang lebih tepat berbanding keputusan diperolehi daripada persamaan regresi individu. Persamaan regresi terbaik untuk menganggarkan kandungan N diperolehi dari data yang digabungkan dan persamaan tersebut ialah $Y=2.7935-0.0014x$ di mana Y ialah kandungan N dan x mewakili bacaan pantulan. Perbezaan purata di antara kandungan N sebenar dan N dijangka daripada data yang terkumpul ialah hanya 0.03%, di mana ianya tidak signifikan apabila diuji secara kaedah t ($P \geq 0.05$, $df=266$). Hasil ujian ini menunjukkan kaedah teknik pantulan menggunakan Spektrometer Mini HR 2000 Gentian Optik Berresolusi Tinggi model Ocean Optic didapati berkesan untuk menentukan kandungan N daun kelapa sawit segar.

ACKNOWLEDGEMENTS

My sincere gratitude goes to Associate Professor Dr. Anuar Abdul Rahim, the Chairman of my Supervisory Committee for his assistance, advice and guidance during the preparation of this thesis. His encouragement and generosity are commendable. I am also grateful to Professor Dr. Mohd Hanafi Musa and Associate Professor Dr. Che Fauziah Ishak, the members of the Supervisory Committee for their invaluable assistance and guidance at the various stages of my research.

I am thankful to the Management of Sime Darby Research Sdn. Bhd., Banting for the support. I also owe the members of sampling team from the Transfer of Technology-Upstream for their cooperation and time. I also like to thank the Chemical Laboratory staff for the analysis with special reference to Encik Chew Tze Chiang, Encik Yusof Daud, Puan Jessy Gomez, Cik Kamsiah Khalit (for the nitrogen analysis), Puan Islinah Amad (for the scanning of the leaf, compilation of the results and for the statistical analysis) and Cik Siti Salwah Salleh from Administration Unit for the printing of this thesis.

My sincere appreciations also go to the Graduate school and staff of Department of Land Management for their cooperation. I also owe the members of my family an incalculable debt for their support and prayers. Best of all, all praises and thanks to ALLAH for the immeasurable guidance, good health and direction.

I certify that a Thesis Examination Committee has met on 13 January 2010 to conduct the final examination of Ahmad Sarwani Haji Adni on his thesis entitled “Determination of Oil Palm Leaf Nitrogen Concentration By Reflectance Spectroscopy” in accordance with Universities and University Colleges Act 1971 and the Constitution of the Universiti Putra Malaysia [P.U.(A) 106] 15 March 1998. The Committee recommends that the student be awarded the Master in Science.

Members of the Thesis Examination Committee were as follows:

Samsuri Abd. Wahid, PhD

Lecturer
Faculty of Agriculture
Universiti Putra Malaysia
(Chairman)

Siva Kumar Balasundram, PhD

Lecturer
Faculty of Agriculture
Universiti Putra Malaysia
(Internal Examiner)

Ahmad Husni Mohd Haniff, PhD

Associate Professor
Faculty of Agriculture
Universiti Putra Malaysia
(Internal Examiner)

Izham Ahmad, PhD

Doctor
Sustainable Development of the Tropical Fruit Industry
International Tropical Fruit Network (TFnet)
Malaysia
(External Examiner)

BUJANG BIN KIM HUAT, PhD

Professor and Deputy Dean
School of Graduate Studies
Universiti Putra Malaysia

Date: 20 May 2010

This thesis was submitted to the senate of Universiti Putra Malaysia and has been accepted as fulfillment of the requirement for the degree of Master in Science. The members of the Supervisory Committee were as follows:

Anuar Abd. Rahim, Ph.D.

Associate Professor
Faculty of Agriculture
Universiti Putra Malaysia
(Chairman)

Che Fauziah Ishak, Ph.D.

Associate Professor
Faculty of Agriculture
Universiti Putra Malaysia
(Member)

Mohd Hanafi Musa, Ph.D.

Professor
Faculty of Agriculture
Universiti Putra Malaysia
(Member)

HASANAH MOHD GHAZALI, PhD

Professor and Dean
School of Graduate Studies
Universiti Putra Malaysia

Date: 10 June 2010

DECLARATION

I hereby declare that the thesis is based on my original work for quotations and citations which have been duly acknowledged. I also declare that it has not been previously or concurrently submitted for any other degree at Universiti Putra Malaysia or other institutions.



AHMAD SARWANI HAJI ADNI

Date : 13 January 2010

TABLE OF CONTENTS

	Page
ABSTRACT	iii
ABSTRAK	v
ACKNOWLEDGEMENTS	vii
APPROVAL	viii
DECLARATION	x
LIST OF TABLES	xiv
LIST OF FIGURES	xvii
LIST OF ABBREVIATIONS	xviii
CHAPTER	
1. INTRODUCTION	1
1.1 Background	2
1.2 Justification	2
1.3 General Objective	6
1.3.1 Specific objectives	7
2. LITERATURE REVIEW	8
2.1 Crop Ecology, Production and Management	8
2.2 Spectrometry	8
2.3 Reflectance Spectroscopy and Reflectance of Leaf Components	14
2.3.1 The Chlorophylls and Its Chemistry	15
2.3.2 Chlorophyll Synthesis	18
2.3.3 Physiological Role of Nitrogen	19
2.3.4 Estimating Nitrogen Concentration in Leaves by Reflectance Measurements	20
2.4 Chlorophyll Meter SPAD-502	24
2.4.1 Theory of SPAD-502	26
2.5 Fiber Optic Spectrometer	27
2.5.1 HR2000 High-Resolution Miniature Fiber Optic Spectrometer	28
2.5.2 HR2000 Specifications	29
3. MATERIALS AND METHODOLOGY	31
3.1 Study 1 - Relationship Between Nitrogen (Kjedahl Method) and Chlorophyll (Chlorophyll Meter SPAD-502) in Fresh Oil Palm Leaf	31
3.1.1 Sampling and Data Collection	31
3.1.2 Indirect Nitrogen Determination	31
3.1.3 Conventional Nitrogen Determination by Kjedadhl Method	32
3.1.4 Statistical Analysis	34

3.2	Study 2 - Determination of Nitrogen Concentration in Fresh Oil Palm Leaf of Different Ages (from 2 to 27 years) by Reflectance Measurement	34
3.2.1	Sampling and Data Collection	34
3.2.2	Sampling of Frond 17	36
3.2.3	Reflection Measurements	37
3.2.4	Conventional Nitrogen Determination by Kjeldahl Method	38
	Refer to sub-section 3.1.3 of Study 1 Relationship Between Nitrogen (Kjeldahl Method) and Chlorophyll (Chlorophyll Meter SPAD-502) in Fresh Oil Palm Leaf.	39
3.2.5	Statistical Analysis	39
3.3	Study 3 - Nitrogen Concentration in Fresh Oil Palm Leaf of One Year Seedling with Different Fertilization Treatments by Reflectance Measurement	39
3.3.1	Sampling and Data Collection	39
3.3.2	Reflection Measurements	41
	Refer to sub-section 3.2.3 of Study 2 - Nitrogen Concentration in Fresh Oil Palm Leaf of Different Ages (from 2 to 25 years) by Reflectance Measurement	
3.3.3	Conventional Nitrogen Determination by Kjeldahl Method	41
	Refer to sub-section 3.1.3 of Study 1 - Relationship Between Nitrogen (Kjeldahl Method) and Chlorophyll (Chlorophyll Meter SPAD-502) in Fresh Oil Palm Leaf	
3.3.4	Statistical Analysis	41
4.	RESULTS AND DISCUSSION	42
4.1	Study 1 - Relationship Between Nitrogen (Kjeldahl Method) and Chlorophyll (Chlorophyll Meter SPAD-502) in Fresh Oil Palm Leaf	42
4.2	Study 2 - Nitrogen Concentration in Fresh Oil Palm Leaf of Different Ages (from 2 to 27 years) by Reflectance Measurement	45
4.2.1	Reflectance Spectra of Fresh Oil Palm Leaf	45
4.3	Selection of Best Wavelength and Best Fit-Line of Fresh Oil Palm Leaf Reflectance Spectra Most Strongly Correlated With Nitrogen Concentration from Kjeldahl Method	48
4.3.1	Study 2 - Selection of Best Wavelength and Best Fit-Line of Fresh Oil Palm Leaf of Different Ages (from 2 to 27 years)	48

4.3.2	Study 3 - Selection of Best Wavelength and Best Fit-Line of Fresh Oil Palm Leaf of One Year Seedling with Different Fertilization Treatments	62
4.4	Study 2 and Study 3 – Prediction of Nitrogen Concentration in Fresh Oil Palm Leaf Using Equation Developed from Reflectance Measurement and Estimation of Precision and Accuracy of Results Against Kjeldahl Method (Validation Step)	70
5.	GENERAL SUMMARY	76
6.	GENERAL CONCLUSION	79
	REFERENCE	82
	APPENDICES	96
	BIODATA OF STUDENT	203