



UNIVERSITI PUTRA MALAYSIA

***OPTIMIZATION TECHNIQUES FOR UNDERFOOT PRESSURE
MEASUREMENT***

OMAR HUSSEIN ALWAN

FK 2017 36



OPTIMIZATION TECHNIQUES FOR UNDERFOOT PRESSURE MEASUREMENT

By

OMAR HUSSEIN ALWAN

**Thesis Submitted to the School of Graduate Studies, Universiti Putra Malaysia,
in Fulfilment of the Requirements of the Degree of Master of Science**

March 2017

All material contained within the thesis, including without limitation text, logos, icons, photographs and all other artwork, is copyright material of Universiti Putra Malaysia unless otherwise stated. Use may be made of any material contained within the thesis for non-commercial purposes from the copyright holder. Commercial use of material may only be made with the express, prior, written permission of Universiti Putra Malaysia.

Copyright © Universiti Putra Malaysia



DEDICATION

*This work is dedicated to.....
My family...*

My country...



Abstract of thesis presented to the Senate of Universiti Putra Malaysia in fulfilment
of the requirement for the degree of Master of Science

OPTIMIZATION TECHNIQUES FOR UNDERFOOT PRESSURE MEASUREMENT

By

OMAR HUSSEIN ALWAN

March 2017

Chairman : Associate Professor Wan Zuha bin Wan Hasan, PhD
Faculty : Engineering

A plantar foot pressure measurement provides important information which can be used as a helpful tool for evaluating patients with foot complaints. Foot plantar pressure systems are devices that are used to measure underfoot pressure. They can be classified into two types: platform systems and insole systems. An insole system is flexible and portable, thereby allowing a wider variety of studies to be carried out with different gait tasks, footwear designs and terrains, while a platform system is fixed and limited to the measurement of underfoot pressure in a static condition, and is usually used for the detection of ulceration. Currently, underfoot pressure ranges have not been effectively established, and there is insufficient information about the minimum and maximum values of underfoot pressure. Likewise, constraints with regard to the implementation of insole systems still persist and are represented by the need to put a spacer on the sensor during measurements, with the problem of sensor parameters changing after the calibration. Therefore, optimization of measurement techniques have been proposed for the implementation of an insole system to measure the underfoot pressure to overcome these drawbacks.

This research was approved by the Human Research Ethics Committee of Universiti Putra Malaysia. The research was aimed at measuring the pressure of 15 points under the foot for 12 samples, divided into two groups (male and female). Two techniques were used to implement the insole system for the effective measurement of underfoot pressure, namely a sensor optimization of the output voltage swing, and the setting of the physical characteristics of a spacer to be placed on the sensors. In addition, a self-calibration technique was employed with the insole system to maintain the sensor measurements and to solve the problem of the changing of the sensor parameters after a period of time following the calibration. The performance of the proposed device after the application of these techniques was tested with an insole system based on the calculation of the body weight. The results showed that the proposed insole plantar pressure device was able to measure the plantar pressure ranges effectively and was

able to calculate the body weight with a measurement error of up to 5.07 % of the real value of the body weight using a weight scale. This was compared to the use of a commercial device (novel EMED system) which calculated the body weight with a measurement error of up to 9.2 %, or the use of the current insole system without any optimization technique, which calculated the body weight with a measurement error of up to 9.06 % from the real value. This work will contribute towards the benefit of the social and industrial sectors in terms of measurement precision.



Abstrak tesis yang dikemukakan kepada Senat Universiti Putra Malaysia
sebagai memenuhi keperluan untuk Ijazah Master Sains

TEKNIK PENGOPTIMUMAN UNTUK PENGUKURAN TEKANAN BAWAH KAKI

Oleh

OMAR HUSSEIN ALWAN

Mac 2017

Pengerusi : Profesor Madya Wan Zuha bin Wan Hasan, PhD
Fakulti : Kejuruteraan

Pengukuran tekanan tapak kaki menyediakan maklumat penting tentang struktur dan fungsi kaki, dan ia adalah alat yang berguna untuk menilai pesakit berpenyakit kaki. Sistem-sistem tekanan tapak kaki adalah alat-alat yang digunakan untuk mengukur tekanan di bawah kaki. Mereka boleh diklasifikasikan kepada dua jenis: sistem-sistem platform dan sistem-sistem lapik dalam kasut. Sistem lapik dalam kasut adalah fleksibel dan mudah alih, dengan itu membolehkan lebih banyak jenis kajian yang boleh dijalankan yang merangkumi tugas-tugas gaya berjalan yang berbeza, reka bentuk kasut dan rupa bumi, manakala sistem platform adalah tetap dan terhad kepada pengukuran tekanan bawah kaki dalam keadaan statik, dan biasanya digunakan untuk mengesan pengulseran. Pada masa ini, julat tekanan bawah kaki belum ditetapkan dengan berkesan, dan tidak terdapat maklumat yang mencukupi mengenai nilai-nilai minimum dan maksimum tekanan bawah kaki. Begitu juga, kekangan berkaitan dengan pelaksanaan sistem lapik dalam kasut masih berterusan dan diwakili oleh keperluan untuk meletakkan peruang di sensor semasa ukuran, serta masalah parameter sensor yang berubah selepas penentukuran. Oleh itu, pengoptimuman teknik pengukuran telah dicadangkan bagi pelaksanaan sistem lapik dalam kasut untuk mengukur tekanan bawah kaki bagi mengatasi kelemahan-kelemahan ini. Kajian ini telah diluluskan oleh Jawatankuasa Etika Penyelidikan Manusia Universiti Putra Malaysia. Kajian ini bertujuan untuk mengukur tekanan 15 titik di bawah kaki bagi 12 sampel yang dibahagikan kepada dua kumpulan (lelaki dan perempuan). Dua teknik telah digunakan untuk melaksanakan sistem lapik dalam kasut bagi pengukuran tekanan bawah kaki yang berkesan, iaitu pengoptimuman sensor untuk ayunan voltan output dan penetapan ciri-ciri fizikal peruang yang akan diletakkan di atas sensor. Di samping itu, teknik penentukuran-sendiri telah digunakan dengan sistem lapik dalam kasut untuk mengekalkan ukuran-ukuran sensor dan untuk menyelesaikan masalah parameter sensor yang berubah selepas satu tempoh masa berikutan penentukuran. Prestasi peranti yang dicadangkan selepas penggunaan teknik-teknik ini telah diuji dengan sistem lapik dalam kasut berdasarkan penghitungan berat badan. Hasil kajian menunjukkan bahawa peranti tekanan tapak

kaki lapik dalam kasut yang dicadangkan itu dapat menghitung julat tekanan tapak kaki dengan berkesan dan menghitung berat badan dengan ralat pengukuran sehingga 5.07% daripada nilai sebenar berat badan menggunakan skala berat badan. Ini telah dibandingkan dengan penggunaan peranti komersial (sistem EMED baru) yang menghitung berat badan dengan ralat pengukuran sehingga 9.2%, atau penggunaan sistem lapik dalam kasut semasa tanpa sebarang teknik pengoptimuman, yang menghitung berat badan dengan ralat pengukuran sehingga 9.06% daripada nilai sebenar. Kajian ini akan memberi manfaat kepada sektor-sektor sosial dan industri dari segi ketepatan pengukuran, penggunaan teknik pengoptimuman dan menyelesaikan masalah parameter sensor yang berubah selepas penentuan.



ACKNOWLEDGEMENTS

In the Name of Allah, Most Gracious, Most Merciful

First and foremost, I would like to thank the Almighty God for the blessing of giving me strength and patience to complete my study.

And I would also like to thank my big family, my mother and my brothers, on their tremendous support and patronage, without them I wouldn't go finish my study. And also I thank my small family, my wife and my daughter, for bearing the rigors of such a long and arduous journey.

I would like to take this opportunity to express my sincere gratitude and appreciation to my supervisor Assoc. Prof. Dr. Wan Zuha b. Wan Hasan for all his guidance and instructions to me during the study. And many thanks for the support given from my co-supervisors Assoc. Prof. Dr. Azura Bt. Che Soh and Dr. Haslina Binti Jaafar

I also would like to thank Universiti Putra Malaysia UPM for accepting me in the Faculty of Engineering.

And last but not the least; I want to thank the hard working student and my friend who has helped me from the beginning of my study Mr. Ahmed Hussein Sabry.

I certify that a Thesis Examination Committee has met on 28 March 2017 to conduct the final examination of Omar Hussein Alwan on his thesis entitled "Optimization Techniques for Underfoot Pressure Measurement" in accordance with the Universities and University Colleges Act 1971 and the Constitution of the Universiti Putra Malaysia [P.U.(A) 106] 15 March 1998. The Committee recommends that the student be awarded the Master of Science.

Members of the Thesis Examination Committee were as follows:

Maryam binti Mohd Isa, PhD


Senior Lecturer
Faculty of Engineering
Universiti Putra Malaysia
(Chairman)

Suhaidi bin Shafie, PhD

Associate Professor
Faculty of Engineering
Universiti Putra Malaysia
(Internal Examiner)

Sawal Hamid Md Ali, PhD

Senior Lecturer
The National University of Malaysia
Malaysia
(External Examiner)



NOR AINI AB. SHUKOR, PhD

Professor and Deputy Dean
School of Graduate Studies
Universiti Putra Malaysia

Date: 2 June 2017

This thesis was submitted to the Senate of Universiti Putra Malaysia and has been accepted as fulfilment of the requirement for the degree of Master of Science. The members of the Supervisory Committee were as follows:

Wan Zuha bin Wan Hasan, PhD

Associate Professor

Faculty of Engineering

Universiti Putra Malaysia

(Chairman)

Azura Bt. Che Soh, PhD

Associate Professor

Faculty of Engineering

Universiti Putra Malaysia

(Member)

Haslina Binti Jaafar, PhD

Senior Lecturer

Faculty of Engineering

Universiti Putra Malaysia

(Member)

ROBIAH BINTI YUNUS, PhD

Professor and Dean

School of Graduate Studies

Universiti Putra Malaysia

Date:

Declaration by graduate student

I hereby confirm that:

- this thesis is my original work;
- quotations, illustrations and citations have been duly referenced;
- this thesis has not been submitted previously or concurrently for any other degree at any other institutions;
- intellectual property from the thesis and copyright of thesis are fully-owned by Universiti Putra Malaysia, as according to the Universiti Putra Malaysia (Research) Rules 2012;
- written permission must be obtained from supervisor and the office of Deputy Vice-Chancellor (Research and Innovation) before thesis is published (in the form of written, printed or in electronic form) including books, journals, modules, proceedings, popular writings, seminar papers, manuscripts, posters, reports, lecture notes, learning modules or any other materials as stated in the Universiti Putra Malaysia (Research) Rules 2012;
- There is no plagiarism or data falsification/fabrication in the thesis, and scholarly integrity is upheld as according to the Universiti Putra Malaysia (Graduate Studies) Rules 2003 (Revision 2012-2013) and the Universiti Putra Malaysia (Research) Rules 2012. The thesis has undergone plagiarism detection software.

Signature: _____ Date: _____

Name and Matric No.: Omar Hussein Alwan, GS40729

Declaration by Member Supervisory Committee

This is to confirm that:

- the research conducted and the writing of this thesis was under our supervision;
- supervision responsibilities as stated in the Universiti Putra Malaysia (Graduate Studies) Rules 2003 (Revision 2012-2013) are adhered to.

Signature: _____
Name of Chairman
of Supervisory Committee: Associate Professor
Dr. Wan Zuha bin Wan Hasan

Signature: _____
Name of Member
of Supervisory Committee: Associate Professor
Dr. Azura Bt. Che Soh

Signature: _____
Name of Member
of Supervisory Committee: Dr. Haslina Binti Jaafar

TABLE OF CONTENTS

	Page
ABSTRACT	i
ABSTRAK	iii
ACKNOWLEDGEMENTS	v
APPROVAL	vi
DECLARATION	viii
LIST OF TABLES	xii
LIST OF FIGURES	xiii
LIST OF ABBREVIATIONS	xv
 CHAPTER	
 1 INTRODUCTION	 1
1.1 Overview	1
1.2 Problem Statement	2
1.3 Research Objectives	3
1.4 Scope and limitation of the study	3
1.5 Research Contributions	4
1.6 Layout of the Thesis	4
 2 LITERATURE REVIEW	 5
2.1 Overview	5
2.2 Trends in Human Motion Measurement	6
2.3 Foot Plantar Pressure Measurements	6
2.3.1 Importance of Foot Plantar Pressure Measurement	6
2.3.2 Common Foot Plantar Pressure Measurement Methods	7
2.4 In-Shoe Sensors	8
2.4.1 Pressure Sensor Calibration	9
2.4.2 Flexi force sensor	11
2.5 Insole Plantar Pressure Measurement Device Requirements	13
2.5.1 In Shoe Sensor Distribution	13
2.5.2 Factors that Affect the Plantar Pressure Measurements	14
2.6 spacer material technique	16
2.7 Self-Calibration for the Pressure Sensor	17
2.7.1 Trends of Self-Calibration for the Pressure Sensor	17
2.7.2 Self-Calibration for Piezo resistive Pressure Sensor	18
2.8 Novel EMED Systems	18
2.9 Chapter Summary	19

3	RESEARCH METHODOLOGY	20
3.1	Overview	20
3.2	Proposed Insole Plantar Pressure Device	20
3.2.1	Optimizing sensor output voltage swing	22
3.2.2	Selection of Supplementary Material for Spacer of Flexi force Pressure Sensor	27
3.2.3	Self-calibration Technique	30
3.3	Underfoot Pressure Measurement Methodology using Proposed IPPD	35
3.4	Calculation of Sample Size	36
3.5	Summary	38
4	RESULTS AND DISCUSSION	39
4.1	Overview	39
4.2	Optimization Measurements Technique with the Proposed IPPD	39
4.2.1	Verification of Measurement Precision for Optimization of Sensor for the voltage output swing	39
4.2.2	Validation of Precision Measurements Based on Calculating the Weight after Applying a Spacer with Sensors	42
4.2.3	Verification of Measurement Precision with Applying Self-Calibration Procedure	43
4.3	Underfoot Pressure Measurements using Proposed IPPD	48
4.4	Comparison of Body Weight Calculations using the Current insole system, Proposed IPPD and Novel EMED System	53
4.5	Increased Sensor sensitivity	55
4.6	Summary	56
5	CONCLUSION AND RECOMMANDATIONS	59
5.1	Conclusion	59
5.2	Recommendation	60
	REFERENCES	61
	APPENDICES	69
	BIODATA OF STUDENT	78
	LIST OF PUBLICATIONS	79

LIST OF TABLES

Table	Page
2.1 In-Shoe Pressure Sensors in the Market	8
4.1 Maximum applied pressure for 15 sensor positions under the foot	40
4.2 Body weight calculation using insole system with two considerations of sensor optimization	41
4.3 Weight calculations using pressure sensors with and without applying a spacer	42
4.4 Sensor output vs. applied force after a period following the calibration	44
4.5 Body weight calculation using insole system	46
4.6 Maximum applied underfoot pressure using IPPD (male group) during walking condition	49
4.7 Maximum applied underfoot pressure using IPPD (female samples) during walking condition	50
4.8 Minimum applied pressure of underfoot using IPPD (female samples) during standing Condition	51
4.9 Minimum applied pressure of underfoot using IPPD (female samples) during standing condition	52
4.10 Comparison between weighing scale, novel EMED system, current insole system, and proposed IPPD in terms of body weight calculations	54
4.11 Summary of the extracted value for different applied techniques	58

LIST OF FIGURES

Figure		Page
2.1	Introductory diagram	5
2.2	Foot plantar system by Zebris Medical GmbH	8
2.3	Current-to-voltage converter circuit diagram	9
2.4	FSR with and without a resin dome	10
2.5	Sensor Output Voltages vs. Force	10
2.6	Sensor Output Conductance (1/Resistance) vs. Force	11
2.7	Sensor driving circuit	12
2.8	Force sensor output voltage vs. applied force (lb)	12
2.9	Sensor distribution of 15 positions	14
2.10	Novel EMED System software	19
3.1	Diagram of proposed insole system (a) Sensor distribution, (b) Sensor driving circuit, (c) Optimization measurement techniques, (d) Underfoot pressure ranges	21
3.2	Flow chart of IPPD Implementation	22
3.3	Flowchart of obtaining maximum applied pressure under the foot	24
3.4	Body weight calculations using an insole system with two sensor situations of optimization	25
3.5	Distribution of sensors under a size-43 foot	26
3.6	Flowchart of weight calculation with and without the use of a spacer	28
3.7	Universal Testing Machine with rubber and plastic spacer with flexi force pressure sensor	29
3.8	Investigation of sensor performance after a period of time	31
3.9	Self-Calibration Procedure	33
3.10	MATLAB Simulink Circuit representing the self-calibration procedure	34

3.11	Standing and walking conditions using the proposed insole plantar pressure device	36
3.12	Standing condition using the novel EMED system	36
3.13	Statistical information on mean value and standard deviation	38
4.1	Proposed Insole Plantar Pressure Device (IPPD)	39
4.2	Percentage of error between calculated value and real value with different spacer thickness	43
4.3	Relationship between the applied force and output voltage after a period of calibration	44
4.4	Self-calibration output	47
4.5	Relationship between output voltage of sensor and applied force after one hour of calibration	48
4.6	Maximum applied pressure with increasing body weight for male group	49
4.7	Maximum applied pressure with increasing body weight for female group	50
4.8	Minimum applied pressure with increasing body weight for male	51
4.9	Minimum applied pressure with increasing body weight for female	52
4.10	Sensitivity of system (mV/kPa)	56

LIST OF ABBREVIATIONS

ANN	Artificial Neural Network
BW	Body Weight
CDPS	Capacitive Differential Pressure Sensor
COP	Center Of Pressure
CPS	Capacitive Pressure Sensor
FSR	Force-Sensing Resistor
IPPD	Insole Plantar Pressure Device
NN	Neural Network
RA	Rheumatoid Arthritis
UTM	Universal Test Machine

CHAPTER 1

INTRODUCTION

1.1 Overview

Feet are very important for human beings to provide stability and support while standing, walking and running. The functions of the foot are summarized as a shock absorber, a mobile adapter, and a rigid lever for the human body [1]. The underfoot pressure provides helpful and important information in research into gait and posture for diagnosing lower limb problems, for footwear designs, sports biomechanics, injury prevention and other applications. This research provides a precise measurement of underfoot pressure, which is important because these measurements can be used by clinicians and doctors for the early detection of the abovementioned problems. Underfoot pressure measurement systems are classified as one of two types: platform systems or insole systems.

The insole plantar pressure device (IPPD) plays a crucial role in the screening, treatment and behaviour modification of patients who are at risk of or are already experiencing a variety of foot and gait problems. The IPPD is used for assessing and improving the quality of life. For example, in sports, it is used to improve and monitor the performance of athletes, to enable people to monitor their health condition, and for them to practise sport postures anywhere at any time [2]. In addition, for patients who are suffering from diabetes, it is used to screen the development of foot ulceration. In terms of gait stability, it has proven to be very helpful in assessing and improving balance among the elderly [3]. It is also used to prevent the risk of ulceration or peripheral neuropathy in patients with diabetes and other sicknesses [4].

The need for an accurate underfoot pressure measurement system is driven by the increasing population of elderly people and the rise in the number of diseases related to a hectic lifestyle. However, underfoot pressure ranges measured through the use of plantar pressure measurement systems in relation to increasing body weight have not been effectively established yet to prove that this cumulative pressure is equal to the body weight of the sample. Commercial devices and current insole system devices can extract these ranges, but with a lot of measurement error in terms of body weight. This error is represented by a difference in the measurement in terms of the calculation of the body weight between the current insole system, current commercial devices and a weight scale. The proposed IPPD was used to obtain the pressure based on increasing the body weight effectively with less measurement error as compared to commercial devices and the current insole system. These pressure values are useful as a reference for researchers to compare their results. To ensure that the pressure range values had been effectively set, the proposed device had to fulfil certain requirements and apply optimization measurement techniques, which included optimization the sensor for voltage output swing, correctly setting the parameters which can affect the measurement represented by the material type and thickness to place on the sensor during the measurement, and applying the self-calibration technique. This research definitely has the potential to find a reference of underfoot pressure range values based

on increasing the body weight with less measurement error in terms of the human body weight.

1.2 Problem Statement

Foot plantar pressure systems are used extensively to measure the pressure pattern underneath the human body so as to provide valuable information about the function of the underfoot pressure. Most underfoot plantar pressure systems are applied to solve problems with regard to gait and posture. Applications of the insole system have been used for many purposes such as to investigate the relationship between increasing body mass and peak/mean pressure [5], or to distinguish the phases of the normal gait cycle in real time [6]. These systems have been used for the reason of extracting the real-time measurements and replicate the shape of the ground reaction force and ankle moment in a stroke patient. Despite different purposes and aims, insole systems lack the ability to measure the body weight for such application to improve the precision of the extracted values [7, 8].

Presently, not many researchers have paid attention to pressure range values and suitable measurement procedures for improving the self-calibration and physical characteristics of the material used for footwear in relation to increasing human body weight. In addition, the current commercial devices which can extract these pressure values are not very precise in their calculation of the body weight. According to an investigation by [9,10] on the impact of increasing body weight (obesity), it was reported that there were disparities in the measurement of the peak plantar pressure in the regions affected by body weight. Furthermore, in addressing the plantar pressure difference during standing and walking in obese and non-obese subjects, it was reported that there was an increase in the value of the peak plantar pressure while walking in all the sections of the foot between the obese and non-obese subjects [9]. Moreover, the studies stated that the significant factor that influenced the foot plantar pressure measurement was the increase in human body weight [11]. The initial evidence indicated that the increase in body weight was a factor that could influence the distribution of plantar pressure in the foot [9, 12]. The researchers suggested that clinicians dealing with foot problems should consider the effect of increased body weight on plantar loading in obese patients [12]. The relationship between increasing body mass and mean as well as peak pressure was dependent upon the plantar region [13]. From the review of these previous studies, it can be seen that the pressure values were not taken into consideration in pressure measurements using the current insole system. Therefore, a significant technique was necessary for attaining points in order to find a suitable pressure range when using the plantar system, particularly the minimum and maximum pressure values that are related to the body mass or the peak plantar pressure during walking and standing conditions. Thus, it was necessary to find the underfoot pressure values based on increasing body weight so that they can be used as a reference by other researchers to compare their results with the extracted results. These pressure values can also be used to differentiate between people without foot complications and people, such as diabetics, with foot complications, as mentioned by [14].

Consistent with [15], Tek-scan, the manufacturer of Flexi force pressure sensors, states that when the footprint of the applied load is larger than the sensing area, it is necessary to use a spacer. This is a piece of material (smaller than the sensing area), that is placed on the sensing area to ensure that the entire load path goes through this area. It is necessary to investigate whether this spacer is required during calibrations and experiments to extract a precise measurement. In addition, the characteristics of this spacer, in terms of the type of material and the thickness need to be explored.

Furthermore, based on the elementary results of experiments with the Flexiforce pressure sensors, and according to the Flexiforce pressure sensor calibration guide, it is recommended that the conditioning of the sensor be repeated after a period of time [16]. It can be concluded that there is a problem if the sensors need to be recalibrated after a particular period to maintain the same measurements.

1.3 Research Objectives

The aim of this research was to develop and implement an IPPD that is able to measure the pressure under the foot effectively. In this work, the underfoot pressure was measured in terms of the body weight. Furthermore, it was also vital to reduce measurement error by obtaining the lowest percentage of error in terms of the body weight calculations with applying optimization techniques. In order to realize the above aims, it was necessary to fulfil the following objectives.

1. To improve sensor sensitivity based on maximum applied pressure and maximum weight involved for this device through optimization measurement.
2. To improve pressure distribution by applying a suitable spacer and specify the characteristics of this material in terms of the type and thickness of the material.
3. To reduce pressure error due to sensor resistance changing by implementing the self-calibration procedure.

1.4 Scope and Limitations of the Study

This research focused on the development of an IPPD for measuring the pressure under the foot. The IPPD was used to find the pressure values limits under the foot for 15 positions that are considered as important points for supporting the body weight. This research managed to achieve its objectives through the participation of 12 samples approved by the Human Research Ethics Committee of Universiti Putra Malaysia for two group of male and female with the group of age named as a young adult. In this project, the researcher also focused on providing a precise measurement of plantar pressure with lower measurement error compared to the current insole system and existing commercial devices, such as the novel system by EMED, which are based on body weight calculations. There were also a few factors that could have influenced and limited the experimental data while the research was being conducted.

First, the IPPD had a limitation in terms of the foot size, where it was set for a Euro foot size of 43, which is equivalent to a UK size of 10. Second, the IPPD was

implemented to measure the pressure under the foot for a specific range of samples with a maximum weight of 100 kg. So, samples that were beyond this range could not be involved with this device. In addition, the experiment was carried out in an indoor environment because the insole system was not wireless but had to have a wired connection.

1.5 Research Contributions

The findings of this research will benefit the society and the industrial sector in view of the fact that plantar pressure measurements play an important role as key health indicators. This research will contribute to the industrial sector by providing an IPPD with optimization measurement techniques that can give precise measurements of underfoot pressure in terms of calculating the body weight as compared with the current insole system and existing commercial devices, such as the novel EMED system. This will help companies to improve measurements by foot plantar pressure devices based on this measurement technique. It is also very important to have a self-calibration technique to ensure that foot plantar pressure devices provide precise measurements and that the same measurements are maintained by the sensors.

This research will also contribute to clinical measurements in providing detailed and useful information concerning underfoot pressure represented by information on the minimum and maximum pressure in relation to increasing body weight. This information can be a helpful guide for the early detection of gait and ulceration problems by clinicians and doctors.

1.6 Layout of the Thesis

This thesis starts with Chapter 1, which presents a general introduction to the subject and the problem statement. It also introduces the aims, objectives, and contribution of the study, and gives a brief summary of the structure of the thesis.

Next, Chapter 2 gives a description of the insole plantar pressure device, the factors that can affect the plantar pressure, and the various self-calibration techniques.

Chapter 3 describes the research methodology that was carried out to achieve the objectives, and discusses the steps that were taken to implement the insole plantar pressure device.

Subsequently, Chapter 4 presents the results and discussion, and verifies the results against the current insole system and a commercial device.

Finally, Chapter 5 gives a summary and the conclusion according to the findings of this research. Suggestions and recommendations for future research in the area are also presented in this final chapter.

REFERENCES

- [1] Curran, S. a, & Dananberg, H. J. (2005). Future of gait analysis: a podiatric medical perspective. *Journal of the American Podiatric Medical Association*, 95(2), 130–42.
- [2] Jung, P. G., Lim, G., & Kong, K. (2013). A mobile motion capture system based on inertial sensors and smart shoes. In *Proceedings - IEEE International Conference on Robotics and Automation* (pp. 692–697).
- [3] Albu-Shamah, A., & Zhan, J. (2012). Towards an Optimizing Model for Older People at Risk of Falls. *2012 ASE/IEEE International Conference on BioMedical Computing (BioMedCom)*, (Social Informatics), 55–58.
- [4] Benbakhti, A. S., Boukhenous, S., Zizoua, C., & Attari, M. (2015). An instrumented shoe for ambulatory prevention of diabetic foot ulceration. *Proceedings of the 2014 4th International Conference on Wireless Mobile Communication and Healthcare - "Transforming Healthcare through Innovations in Mobile and Wireless Technologies", MOBIHEALTH 2014*, 43–46.
- [5] T. E., T. L.F., A. U. R., H. T.P., G. E., and M. L.C., "Static and frequency domain analysis of plantar pressure distribution in obese and non-obese subjects," *Journal of Bodywork and Movement Therapies*, vol. 10, no. 2. pp. 127–133, 2006.
- [6] W. Srivises, I. Nilkhamhang, and K. Tungpimolrut, "Design of a smart shoe for reliable gait analysis using state transition theory," in *2012 9th International Conference on Electrical Engineering/Electronics, Computer, Telecommunications and Information Technology*, 2012, pp. 1–4.
- [7] A. M. Tan, F. K. Fuss, Y. Weizman, Y. Woudstra, and O. Troynikov, "Design of Low Cost Smart Insole for Real Time Measurement of Plantar Pressure," *Procedia Technol.*, vol. 20, no. May 2016, pp. 117–122, 2015.
- [8] A. M. Howell, T. Kobayashi, T. R. Chou, W. Daly, M. Orendurff, and S. J. M. Bamberg, "A laboratory insole for analysis of sensor placement to determine ground reaction force and ankle moment in patients with stroke," in *2012 Annual International Conference of the IEEE Engineering in Medicine and Biology Society*, 2012, pp. 6394–6397
- [9] A. P. Hills, E. M. Hennig, M. McDonald, and O. Bar-Or, "Plantar pressure differences between obese and non-obese adults: a biomechanical analysis." *Int. J. Obes. Relat. Metab. Disord.*, vol. 25, no. 11, pp. 1674–9, 2001.
- [10] M. Birtane and H. Tuna, "The evaluation of plantar pressure distribution in obese and non-obese adults," *Clin. Biomech.*, vol. 19, no. 10, pp. 1055–1059, 2004.

- [11] T. E., T. L.F., A. U. R., H. T.P., G. E., and M. L.C., "Static and frequency domain analysis of plantar pressure distribution in obese and non-obese subjects," *Journal of Bodywork and Movement Therapies*, vol. 10, no. 2. pp. 127–133, 2006.
- [12] P. A. Butterworth, D. M. Urquhart, K. B. Landorf, A. E. Wluka, F. M. Cicuttini, and H. B. Menz, "Foot posture, range of motion and plantar pressure characteristics in obese and non-obese individuals," *Gait Posture*, vol. 41, no. 2, pp. 465–469, 2015.
- [13] J. B. Arnold, R. Causby, and S. Jones, "The impact of increasing body mass on peak and mean plantar pressure in asymptomatic adult subjects during walking," *Diabet. Foot Ankle*, vol. 1, 2010.
- [14] U. Flexi, F. Sensors, and M. Chaskar, "Development of Planter Foot Pressure Distribution System," vol. 108, no. 9, pp. 73–79, 2009.
- [15] Tekscan. (2009). Flexiforce User Manual. Retrieved July 1, 2016, from http://www.phidgets.com/documentation/Phidgets/3100_0_FlexiforceUserManual.pdf
- [16] Tekscan, "How do I calibrate my FlexiForce Sensor?," 2013. [Online]. Available: <https://www.tekscan.com/support/faqs/how-do-i-calibrate-my-flexiforce-sensor>. [Accessed: 01-Jul-2016].
- [17] S. a Curran and H. J. Dananberg, "Future of gait analysis: a podiatric medical perspective.," *J. Am. Podiatr. Med. Assoc.*, vol. 95, no. 2, pp. 130–42, 2005.
- [18] F. Yong, G. Yunjian, and S. Quanjun, "A human identification method based on dynamic plantar pressure distribution," in *Information and Automation (ICIA), 2011 IEEE International Conference on*, 2011, pp. 329–332.
- [19] M. Jain, U. Tiwari, and M. Gupta, "Mobile charger via walk," in *2011 International Conference on Multimedia, Signal Processing and Communication Technologies, IMPACT 2011*, 2011, pp. 149–152.
- [20] D. C. Billing, C. R. Nagarajah, J. P. Hayes, and J. Baker, "Predicting ground reaction forces in running using micro-sensors and neural networks," *Sport. Eng.*, vol. 9, no. 1, pp. 15–27, 2006.
- [21] K. Aminian and B. Najafi, "Capturing human motion using body-fixed sensors: Outdoor measurement and clinical applications," *Comput. Animat. Virtual Worlds*, vol. 15, no. 2, pp. 79–94, 2004.
- [22] Y. S. A. Mustufa, J. Barton, B. O'Flynn, R. Davies, P. McCullagh, and H. Zheng, "Design of a smart insole for ambulatory assessment of gait," *2015 IEEE 12th Int. Conf. Wearable Implant. Body Sens. Networks, BSN 2015*, 2015.
- [23] S. R. Simon, "Quantification of human motion: gait analysis - benefits and

- limitations to its application to clinical problems,” *J. Biomech.*, vol. 37, no. 12, pp. 1869–1880, 2004.
- [24] R. Begg, R. Best, L. Dell’Oro, and S. Taylor, “Minimum foot clearance during walking: Strategies for the minimisation of trip-related falls,” *Gait Posture*, vol. 25, no. 2, pp. 191–198, 2007.
 - [25] W. Donkrajang, N. Watthanawisuth, J. P. Mensing, and T. Kerdcharoen, “Development of a wireless electronic shoe for walking abnormalities detection,” in *5th 2012 Biomedical Engineering International Conference, BMEiCON 2012*, 2012.
 - [26] S. K. Sabut, R. Kumar, and M. Mahadevappa, “Design of an insole embedded foot pressure sensor controlled FES system for foot drop in stroke patients,” in *International Conference on Systems in Medicine and Biology, ICSMB 2010 - Proceedings*, 2010, pp. 237–241.
 - [27] T. Yanbo, Q. Huihuan, C. Meng, S. Xin, and X. Yangsheng, “A Real-time intelligent shoe system for fall detection,” in *Robotics and Biomimetics (ROBIO), 2011 IEEE International Conference on*, 2011, pp. 2253–2258.
 - [28] H. Sobral, A. Vieira, J. P. Ferreira, P. Ferreira, S. Cruz, M. Crisostomo, and A. P. Coimbra, “Human gait analysis using instrumented shoes,” *2015 IEEE 4th Port. Meet. Bioeng.*, vol. 14, no. February, p. 1, 2015.
 - [29] J. Bryzek, S. Roundy, B. Bircumshaw, C. Chung, K. Castellino, J. R. Stetter, and M. Vestel, “Marvelous MEMs,” *IEEE Circuits Devices Mag.*, vol. 22, no. 2, pp. 8–28, 2006.
 - [30] E. Jovanov, A. Milenkovic, C. Otto, and P. C. de Groen, “A wireless body area network of intelligent motion sensors for computer assisted physical rehabilitation,” *J. Neuroeng. Rehabil.*, vol. 2, no. 1, p. 6, 2005.
 - [31] M. J. Mueller, “Application of plantar pressure assessment in footwear and insert design,” *J. Orthop. Sports Phys. Ther.*, vol. 29, no. 12, pp. 747–755, 1999.
 - [32] A. Jaume-I-Cap??, P. Martinez-Bueso, B. Moya-Alcover, and J. Varona, “Interactive rehabilitation system for improvement of balance therapies in people with cerebral palsy,” *IEEE Trans. Neural Syst. Rehabil. Eng.*, vol. 22, no. 2, pp. 419–427, 2014.
 - [33] M. Eid, U. Saad, and U. Afzal, “A real time vibrotactile biofeedback system for optimizing athlete training,” *Haptic Audio Vis. Environ. Games (HAVE), 2013 IEEE Int. Symp.*, pp. 1–6, 2013.
 - [34] L. Ren, D. Li, C. Liu, Y. Yang, Y. Qian, S. Yang, F. Pu, and H. Niu, “Design of in-shoe plantar pressure monitoring system for daily activity exercise stress assessment,” *Biomed. Eng. Informatics (BMEI), 2011 4th Int. Conf.*, vol. 3, pp. 1367–1370, 2011.

- [35] S. Wild, G. Roglic, A. Green, R. Sicree, and H. King, "Global Prevalence of Diabetes," *Diabetes Care*, vol. 27, no. 5, pp. 1047–1053, 2004.
- [36] J. R. Mackey and B. L. Davis, "Simultaneous shear and pressure sensor array for assessing pressure and shear at foot/ground interface.," *J. Biomech.*, vol. 39, no. 15, pp. 2893–7, 2006.
- [37] R. Best and R. Begg, "Overview of movement analysis and gait features," *Libr. en Google Play Comput. Intell. Mov. Sci. neutral networks other Emerg. Tech.*, pp. 1–69, 2006.
- [38] D. J. Margolis, J. Knauss, W. Bilker, and M. Baumgarten, "Medical conditions as risk factors for pressure ulcers in an outpatient setting," *Age Ageing*, vol. 32, no. 3, pp. 259–264, 2003.
- [39] M. Chen, B. Huang, and Y. Xu, "Intelligent shoes for abnormal gait detection," in *Proceedings - IEEE International Conference on Robotics and Automation*, 2008, pp. 2019–2024.
- [40] M. Bouchard and V. S. Mosca, "Flatfoot Deformity in Children and Adolescents : Surgical Indications and Management," *J. Am. Acad. Orthop. Surg.*, vol. 22, no. 10, pp. 623–632, 2014.
- [41] B. M. Trindade, Y. Ono, E. D. Lemaire, and I. Almohimeed, "Development of a wearable ultrasonic sensor and method for continuous monitoring of mechanical properties of plantar soft tissue for diabetic patients," in *IEEE International Ultrasonics Symposium, IUS*, 2014, pp. 2112–2115.
- [42] B. Hodgson, L. Tis, S. Cobb, S. McCarthy, and E. Higbie, "The effect of 2 different custom-molded corrective orthotics on plantar pressure.," *J. Sport Rehabil.*, vol. 15, no. 1, pp. 33–44, 2006.
- [43] R. C. Lawrence, C. G. Helmick, F. C. Arnett, R. A. Deyo, D. T. Felson, E. H. Giannini, S. P. Heyse, R. Hirsch, M. C. Hochberg, G. G. Hunder, M. H. Liang, S. R. Pillemer, V. D. Steen, and F. Wolfe, "Estimates of the prevalence of arthritis and selected musculoskeletal disorders in the United States," *Arthritis Rheum.*, vol. 41, no. 5, pp. 778–799, 1998.
- [44] S. A. Bus, "Priorities in offloading the diabetic foot," *Diabetes. Metab. Res. Rev.*, vol. 28, no. S1, pp. 54–59, Feb. 2012.
- [45] S. A. Bus, R. Haspels, and T. E. Busch-Westbroek, "Evaluation and optimization of therapeutic footwear for neuropathic diabetic foot patients using in-shoe plantar pressure analysis," *Diabetes Care*, vol. 34, no. 7, pp. 1595–1600, 2011.
- [46] P. S. Dyer and S. J. M. Bamberg, "Instrumented insole vs. force plate: A comparison of center of plantar pressure," in *Proceedings of the Annual International Conference of the IEEE Engineering in Medicine and Biology Society, EMBS*, 2011, pp. 6805–6809.

- [47] A. Perrier, N. Vuillerme, V. Luboz, M. Bucki, F. Cannard, B. Diot, D. Colin, D. Rin, J. P. Bourg, and Y. Payan, "Smart Diabetic Socks: Embedded device for diabetic foot prevention," *IRBM*, vol. 35, no. 2, pp. 72–76, 2014.
- [48] M. L. Zequera, S. E. Solomonidis, F. Vega, and L. M. Rondon, "Study of the plantar pressure distribution on the sole of the foot of normal and diabetic subjects in the early stages by using a hydrocell pressure sensor," in *Proceedings of the 25th Annual International Conference of the IEEE Engineering in Medicine and Biology Society (IEEE Cat. No.03CH37439)*, 2003, vol. 2, pp. 1874–1877.
- [49] Abdul Razak, A.H. et al., 2012. Foot plantar pressure measurement system: A review. *Sensors (Switzerland)*, 12(7), pp.9884–9912.
- [50] "Novel Quality in Measurement." [Online]. Available: <http://www.novel.de/novelcontent/pedar>. [Accessed: 01-Jul-2016].
- [51] "Pressure Mapping, Force Measurement & Tactile Sensors," *F-Scan System*. [Online]. Available: <https://www.tekscan.com/products-solutions/systems/f-scan-system?tab=sensors>. [Accessed: 01-Jul-2016].
- [52] J. A. Flórez and A. Velásquez, "Calibration of force sensing resistors (fsr) for static and dynamic applications," in *2010 IEEE ANDESCON Conference Proceedings, ANDESCON 2010*, 2010.
- [53] F. M. & T. S. Pressure Mapping, "Flexiforce pressure sensor." [Online]. Available: <https://www.tekscan.com/flexiforce-load-force-sensors-and-systems>. [Accessed: 01-Jul-2016].
- [54] O. Mazumder, A. S. Kundu, and S. Bhaumik, "Development of wireless insole foot pressure data acquisition device," *Commun. Devices Intell. Syst. (CODIS), 2012 Int. Conf.*, pp. 302–305, 2012.
- [55] C. Zhu, A. Song, and Y. Ma, "A wireless gait measuring instrument based on plantar pressure measuring," in *Proceedings - IEEE 2011 10th International Conference on Electronic Measurement and Instruments, ICEMI 2011*, 2011, vol. 1, pp. 146–150.
- [56] O. Hussein, W. Hasan, and A. Soh, "An Accurate Setting for Remapping Process of Foot Plantar Pressure," *Adv. Sci.*, 2015.
- [57] U. Manupibul, W. Charoensuk, and P. Kaimuk, "Design and development of SMART insole system for plantar pressure measurement in imbalance human body and heavy activities," in *BMEiCON 2014 - 7th Biomedical Engineering International Conference*, 2015.
- [58] T. . W. Y. . L. Q. . F. D. D. . c T. X. . Shu L.a Hua, "In-shoe plantar pressure measurement and analysis system based on fabric pressure sensing array," *IEEE Trans. Inf. Technol. Biomed.*, vol. 14, no. 3, pp. 767–775, 2010

- [59] R. Karkokli and K. M. V. McConville, "Design and development of a cost effective plantar pressure distribution analysis system for the dynamically moving feet.," *Conf. Proc. IEEE Eng. Med. Biol. Soc.*, vol. 1, pp. 6008–11, 2006.
- [60] U. Manupibul, W. Charoensuk, and P. Kaimuk, "Design and development of SMART insole system for plantar pressure measurement in imbalance human body and heavy activities," in *BMEiCON 2014 - 7th Biomedical Engineering International Conference*, 2015.
- [61] M. D. C. Dos Reis, F. A. Soares, A. F. Da Rocha, J. L. A. Carvalho, and S. S. F. R. Rodrigues, "Insole with pressure control and tissue neoformation induction systems for diabetic foot," in *2010 Annual International Conference of the IEEE Engineering in Medicine and Biology Society, EMBC'10*, 2010, pp. 5748–5751.
- [62] A. M. Howell, T. Kobayashi, T. R. Chou, W. Daly, M. Orendurff, and S. J. M. Bamberg, "A laboratory insole for analysis of sensor placement to determine ground reaction force and ankle moment in patients with stroke," in *2012 Annual International Conference of the IEEE Engineering in Medicine and Biology Society*, 2012, pp. 6394–6397.
- [63] S. M. M. De Rossi, T. Lenzi, N. Vitiello, M. Donati, A. Persichetti, F. Giovacchini, F. Vecchi, and M. C. Carrozza, "Development of an in-shoe pressure-sensitive device for gait analysis," in *Proceedings of the Annual International Conference of the IEEE Engineering in Medicine and Biology Society, EMBS*, 2011, pp. 5637–5640.
- [64] C. Giacomozzi, "Appropriateness of plantar pressure measurement devices: A comparative technical assessment," *Gait Posture*, vol. 32, no. 1, pp. 141–144, 2010.
- [65] M. Monteiro, R. Gabriel, J. Aranha, M. N. e Castro, M. Sousa, and M. Moreira, "Influence of obesity and sarcopenic obesity on plantar pressure of postmenopausal women," *Clin. Biomech.*, vol. 25, no. 5, pp. 461–467, 2010.
- [66] H. Deepashini, B. Omar, A. Paungmali, N. Amaramalar, H. Ohnmar, and J. Leonard, "An insight into the plantar pressure distribution of the foot in clinical practice: Narrative review," *Polish Annals of Medicine*, vol. 21, no. 1, pp. 51–56, 2014.
- [67] M. J. Hessert, M. Vyas, J. Leach, K. Hu, L. A. Lipsitz, and V. Novak, "Foot pressure distribution during walking in young and old adults.," *BMC Geriatr.*, vol. 5, p. 8, 2005.
- [68] K. Bosch, A. Nagel, L. Weigend, and D. Rosenbaum, "From 'first' to 'last' steps in life - Pressure patterns of three generations," *Clin. Biomech.*, vol. 24, no. 8, pp. 676–681, 2009.
- [69] D. F. Murphy, B. D. Beynnon, J. D. Michelson, and P. M. Vacek, "Efficacy of

plantar loading parameters during gait in terms of reliability, variability, effect of gender and relationship between contact area and plantar pressure.," *Foot ankle Int. / Am. Orthop. Foot Ankle Soc. [and] Swiss Foot Ankle Soc.*, vol. 26, no. 2, pp. 171–179, 2005.

- [70] A. B. Putti, G. P. Arnold, and R. J. Abboud, "Foot pressure differences in men and women," *Foot Ankle Surg.*, vol. 16, no. 1, pp. 21–24, 2010.
- [71] R. Periyasamy, A. Mishra, S. Anand, and A. C. Ammini, "Preliminary investigation of foot pressure distribution variation in men and women adults while standing," *Foot*, vol. 21, no. 3, pp. 142–148, 2011.
- [72] M. Ko, L. Hughes, and H. Lewis, "Walking speed and peak plantar pressure distribution during barefoot walking in persons with diabetes," *Physiother. Res. Int.*, vol. 17, no. 1, pp. 29–35, 2012.
- [73] M. Maetzler, T. Bochdansky, and R. J. Abboud, "Normal pressure values and repeatability of the Emed?? ST2 system," *Gait Posture*, vol. 32, no. 3, pp. 391–394, 2010.
- [74] A. J. Taylor, H. B. Menz, and A.-M. Keenan, "The influence of walking speed on plantar pressure measurements using the two-step gait initiation protocol," *Foot*, vol. 14, pp. 49–55, 2004.
- [75] J. M. Burnfield, C. D. Few, O. S. Mohamed, and J. Perry, "The influence of walking speed and footwear on plantar pressures in older adults," *Clin. Biomech.*, vol. 19, no. 1, pp. 78–84, 2004.
- [76] O. Hussein, W. Z. W. Hasan, A. C. Soh, and H. Jafaar, "Fundamental references over insole plantar pressure in terms of human body weight percentage," in *2015 IEEE Regional Symposium on Micro and Nanoelectronics (RSM)*, 2015, pp. 1–4.
- [77] Wertsch, J. J., Webster, J. G., & Tompkins, W. J. (1992). A portable insole plantar pressure measurement system. *Journal of rehabilitation research and development*, 29(1), 13.
- [78] Meyer, J., Arnrich, B., Schumm, J., & Troster, G. (2010). Design and modeling of a textile pressure sensor for sitting posture classification. *IEEE Sensors Journal*, 10(8), 1391-1398.
- [79] Tung, T. T., Robert, C., Castro, M., Feller, J. F., Kim, T. Y., & Suh, K. S. (2016). Enhancing the sensitivity of graphene/polyurethane nanocomposite flexible piezo-resistive pressure sensors with magnetite nano-spacers. *Carbon*, 108, 450-460.
- [80] Holleczeck, T., Rüegg, A., Harms, H., & Tröster, G. (2010, November). Textile pressure sensors for sports applications. In *Sensors*, 2010 IEEE (pp. 732-737). IEEE.

- [81] Meyer, J., Arnrich, B., Schumm, J., & Troster, G. (2010). Design and modeling of a textile pressure sensor for sitting posture classification. *IEEE Sensors Journal*, 10(8), 1391-1398.
- [82] Pekárek, J., Ficek, R., Vrba, R., & Magát, M. (2009, September). Carbon nanostructures used in capacitive sensors as the surface increase element. In *Design and Technology of Electronics Packages,(SIITME) 2009 15th International Symposium for* (pp. 323-326). IEEE.
- [83] Paul Muller., “Validyne Engineering Blog.” [Online]. Available: <http://validyne.com/blog/basics-validyne-pressure-sensor-calibration/>. [Accessed: 01-Jul-2016].
- [84] K. Kasten, N. Kordas, H. Kappert, and W. Mokwa, “Capacitive pressure sensor with monolithically integrated CMOS readout circuit for high temperature applications,” in *Sensors and Actuators, A: Physical*, 2002, vol. 97–98, pp. 83–87.
- [85] S. K. Rath, J. C. Patra, and A. C. Kot, “An intelligent pressure sensor with self-calibration capability using artificial neural networks,” in *SMC 2000 Conference Proceedings. 2000 IEEE International Conference on Systems, Man and Cybernetics. “Cybernetics Evolving to Systems, Humans, Organizations, and their Complex Interactions” (Cat. No.00CH37166)*, 2000, vol. 4, pp. 2563–2568.
- [86] J. C. Patra, V. Gopalkrishnan, A. Ee Luang, and A. Das, “Neural network-based self-calibration/compensation of sensors operating in harsh environments [smart pressure sensor example],” *Sensors*, 2004. *Proc. IEEE*, vol. 1, pp. 425–428 vol.1, 2004.
- [87] P. Yameogo, U. Heiba, M. Al Bahri, and P. Pons, “Self calibrating pressure sensor for biomedical applications,” in *Proceedings of IEEE Sensors*, 2009, pp. 691–694.
- [88] C. Doyle, D. Riordan, and J. Walsh, “Automated Bias-removal Resistance Measurement Circuit for Precision On-Site Temperature Calibration Exploitation of State Changes of Materials,” pp. 229–234, 2015.
- [89] The emed®-systems, “No Title.” [Online]. Available: <http://www.novel.de/novelcontent/emed>. [Accessed: 01-Jul-2016].
- [90] I. Str, G. Avenue, G. Britain, and A. Road, “® Systems Pedography,” 2011.
- [91] The emed®-systems, “novel.de.” [Online]. Available: <http://www.novel.de/novelcontent/emed>. [Accessed: 01-Jul-2016].
- [92] F. Semiconductor, “Analog to Digital Converter Resolution Extension Using a Pressure Sensor,” pp. 1–4, 2005.